

INTOXILYZER® 8000 INITIAL INSPECTION

Serial Number: 80-005946

1. Power On and Warm-up:
 - a. Power On
 - b. 5 min. Countdown from Initial Warm-up
 - c. 1 min. Countdown to Begin a Test
 - d. Forms Load Complete

2. Setup and Checks:
 - a. Level 1:
 - 1) Passwords: Field Inspector (Level 2) Lab Personnel (Level 3)
 - 2) Gas Cylinder: Gas Cylinder Setup (Level 1,S)
 - 3) Printer: Printer Setup (Level 3,S,P)
 Internal Printer Paper Feed
 Internal Printer Test (Level 1,P)
 External Printer Test (Level 1,P)
 - 4) Barcode Reader (Level 1,B): Operator ID
 ND License

 - b. Level 2:
 - 1) Continuous Purge (Level 2,A)
 - 2) Diagnostic (Level 2,D)
 - 3) Prelim Data Entry (Level 2,E)
 - 4) Ser Num/Version Info (Level 2,V): Ser Num 80-005946
Version 8164.13

 - c. Level 3:
 - 1) DVM Monitor (Level 3,D,D): 3: 12586 s: 1 C: 47.00
9: 13366 s: 3 B: 44
(RFI indicator with keyed radio) 1

 - 2) Function (F) Keys: F5 – Tone
 F6 – Green LED
 F7 – Red LED
 F9 – Fan
 F10 – Pump
 F11 – Solenoid

 - 3) Atmospheric Monitor (Level 3,D,A): Atmo Sensor, P: 956 CF: 1.05
 - 4) Flow Monitor (Level 3,D,F): Flow Sensor, R: 194 F: 0.000 L/s

5) Temperature Monitor (Level 3,D,T): C: 47.0 A: 48.5 I: 0.0
B: 45.0 V: 0.0

- 6) Tank Monitor (Level 3,D,G): P (gauge): 2990 P (display): 979
7) General Setup (Level 3,S,G)
8) Configure Start Test (Level 3,S,T)
9) Network (Level 3,S,N)

3. Tests:

a. Configure Standard for Wet Bath Simulator (Level 1,S):

1) Zero Alcohol Test (Level 1,C):

Sim Ser #: G-10329 Lot #: --- Exp Date: N/A
Known Value: 0.000 AC
Results: 0.000 AC 0.000 AC 0.000 AC

2) Low AC Test (Level 1,C):

Sim Ser #: DR3803 Lot #: 13070 Exp Date: 03/06/2015
Known Value: 0.020 AC
Results: 0.020 AC 0.020 AC 0.020 AC

3) Linearity Test (Level 1,C):

Sim Ser #: DR5143 Lot #: 13080 Exp Date: 03/19/2015
Known Value: 0.300 AC
Results: 0.294 AC 0.294 AC 0.294 AC

b. Configure Dry Gas Standard (Level 1,S):

1) Interferent Detected Tests (Level 1,B):

1. 0.10 AC + 0.05% Acetone

Sim Ser #: DR 3377 Lot #: 1051 Exp Date: 09 Apr 14
Known Value: 0.100 AC + 0.050% acetone
 Display reads "Interferent Detect"

2. Acetone

Sim Ser #: DR 3375 Lot #: ACE1 Exp Date: 27 Aug 15
Known Value: 0.05% acetone
 Display reads "Interferent Detect"

3. Methanol

Sim Ser #: DR 3454 Lot #: MET1 Exp Date: 27 Aug 15
Known Value: 0.10% Methanol
 Display reads "Interferent Detect"

4. Isopropanol
Sim Ser #: DR5131 Lot #: IPA 1 Exp Date: 27 Aug 15
Known Value: 0.1070 Isopropanol
 Display reads "Interferent Detect"

- 2) Ambient Fail (Level 1,B): Display and test record are correct
- 3) Invalid Sample (Level 1,B): Display and test record are correct
- 4) Subject Test Timeout (Level 1,B): Display and test record are correct.
- 5) Sequence Aborted (Level 1,B): Display and test record are correct
- 6) Control Outside Tolerance (CMS mode): Display and test record are correct
- 7) RFI Detected Test (CMS mode): Display and test record are correct
- 8) Tank Pressure Below Min (CMS mode): Display reads correct
- 9) Subject Refused (CMS mode): Display and test record are correct
- 10) Difference Too Great (CMS mode): Display and test record are correct
- 11) Stability Test (Level 3,T,S):
Average 0.0800 Std Dev 0.0006 Rel Std Dev 0.82%

Maintenance/Repairs/Notes: Initial inspection

Deb Kashur
Field Inspector's Signature

13 NOV 13
Date

Charles E. Eder
Reviewed by

13 NOV 13
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 08:49

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMNQRSTUWX

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup
Data Entry Mode: Disabled
Start Test Sequence: ABABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On MenFull? No
of Print Copies? 1
Select Std (D/W/I)? Wet
Standard Value? 0.100
Standard Solution #?
Sim Serial #?
Oper No? 777777

Flow Cal. Date: 11/01/2013
Slope 628
Intercept -526243

IR Calibration Date: 11/01/2013
3um 9um
0th Coef(*100): -20566 -12505
1st Coef(*100): 270596 135247
2nd Coef(*100): 4183 1305
H2O adj(mg/l*10K): 725 395

***** Test End *****

Deb Keshur
Operator Signature
DEB KASHUR

Remarks:
Print test
Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 08:51

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Disabled
Start Test Sequence: ABABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? No
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 16512080A2
Standard Cyl #? 22
Standard Expiration? 08/01/2014
Oper No? 777777

Flow Cal. Date: 11/01/2013
Slope 628
Intercept -526243

IR Calibration Date: 11/01/2013

	3um	9um
0th Coef(*100):	-20566	-12905
1st Coef(*100):	270596	135247
2nd Coef(*100):	4183	1305
H2O adj(mg/l*10k):	725	395

***** Printer Test End *****



Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 08:51

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

^{DK}
Diagnostic test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 09:23

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:24
02 Std. Sol.	0.000	09:25
03 Room Air	0.000	09:25
04 Std. Sol.	0.000	09:26
05 Room Air	0.000	09:26
06 Std. Sol.	0.000	09:27
07 Room Air	0.000	09:28

08 Sim Temp = 34.0°C

Simul Ser No = G10329

Std Sol No = 0000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Zero alcohol 0.000 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 09:31

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:31
02 Std. Sol.	0.020	09:32
03 Room Air	0.000	09:33
04 Std. Sol.	0.020	09:33
05 Room Air	0.000	09:34
06 Std. Sol.	0.020	09:35
07 Room Air	0.000	09:35

08 Sim Temp = 34.0°C

Simul Ser No = DR3803

Std Sol No = 13070

County = 08

Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Low AC 0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 09:43

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:43
02 Std. Sol.	0.294	09:44
03 Room Air	0.000	09:45
04 Std. Sol.	0.294	09:45
05 Room Air	0.000	09:46
06 Std. Sol.	0.294	09:47
07 Room Air	0.000	09:47

08 Sim Temp = 34.0°C

Simul Ser No = DR5143
Std Sol No = 13080
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.300 AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 10:36

Test	AC	Time
01 Room Air	0.000	10:36
02 *Subject Test	INT*	10:37
03 Room Air	0.000	10:37

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent test

0.100 AC + 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 10:38

Test	AC	Time
01 Room Air	0.000	10:39
02 *Subject Test	INT*	10:39
03 Room Air	0.000	10:40

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent test 0.05% acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 10:40

Test	AC	Time
01 Room Air	0.000	10:41
02 *Subject Test	INT*	10:41
03 Room Air	0.000	10:42

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent test 0.10% methanol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 10:42

Test	AC	Time
01 Room Air	0.000	10:42
02 *Subject Test	INT*	10:43
03 Room Air	0.000	10:44

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Interferent test

0.10% Isopropanol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 10:45

Test	AC	Time
01 Room Air	AMB*	10:46
02 Room Air	0.000	10:46

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

ambient fail test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 10:48

Test	AC	Time
01 Room Air	0.000	10:48
02 Invalid Sample	X.XXX	10:49
03 Room Air	0.000	10:50

*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Invalid sample test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 10:51

Test	AC	Time
01 Room Air	0.000	10:52
02 *Subject Test	0.000*	10:55
03 Room Air	0.000	10:55

No RFI Detected

*Deficient Sample - Value Printed was
Highest Obtained

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:



Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 10:56

Test	AC	Time
01 Room Air	ABT*	10:56
02 Room Air	0.000	10:57

*Sequence Aborted

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Sequence Aborted Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer	Alcohol Analyzer
North Dakota Model 8000	SN 80-005946
Location = TOXL	8164.13.00 06/09
11/13/2013	10:58

Test	AC	Time
01 Diagnostic	OK	10:59
02 Room Air	0.000	10:59
03 Subject Test 1	0.000	11:00
04 Room Air	0.000	11:02
05 Std. Gas	0.000*	11:03
06 Room Air	0.000	11:04

*Cal Check Out of Tolerance

Sub Name = DISCOVER, THE SPIRIT
 Sub DOB = 02/01/1992
 Sub Sex = Female Weight = 150
 Test = DUI Cit = NA
 Dr. Lic. = ND/DIS921456
 Lot No = 16512080A2
 Cyl No = 22
 Expiration Date = 08/01/2014
 County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
 DEB KASHUR

Remarks:

Control outside tolerance test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 11:05

Test	AC	Time
01 Diagnostic	OK	11:07
02 Room Air	RFI*	11:07
03 Room Air	0.000	11:07

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI detected test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 11:11

Test	AC	Time
01 Diagnostic	OK	11:11
02 Room Air	0.000	11:12
03 *Subject Test	REF*	11:12
04 Room Air	0.000	11:12

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DEB KASHUR

Remarks:

Subject refused test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 11:15

Test	AC	Time
01 Diagnostic	OK	11:16
02 Room Air	0.000	11:16
03 Subject Test 1	0.052*	11:16
04 Room Air	0.000	11:19
05 Std. Gas	0.080	11:20
06 Room Air	0.000	11:21
07 Subject Test 2	0.000*	11:21
08 Room Air	0.000	11:23

*Difference Too Great

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 16512080A2
Cyl No = 22
Expiration Date = 08/01/2014
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Difference too great test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005946
Location = TOXL 8164.13.00 06/09
11/13/2013 11:25

Test	AC	Time
01 Room Air	0.000	11:25
02 Std. Gas	0.081	11:26
03 Room Air	0.000	11:26
04 Std. Gas	0.080	11:27
05 Room Air	0.000	11:27
06 Std. Gas	0.080	11:27
07 Room Air	0.000	11:28
08 Std. Gas	0.078	11:28
09 Room Air	0.000	11:29
10 Std. Gas	0.080	11:29
11 Room Air	0.000	11:30
12 Std. Gas	0.081	11:30
13 Room Air	0.000	11:31
14 Std. Gas	0.080	11:31
15 Room Air	0.000	11:31
16 Std. Gas	0.080	11:32
17 Room Air	0.000	11:32
18 Std. Gas	0.080	11:33
19 Room Air	0.000	11:33
20 Std. Gas	0.080	11:34
21 Room Air	0.000	11:34
22 Std. Gas	0.081	11:34
23 Room Air	0.000	11:35
24 Std. Gas	0.080	11:35
25 Room Air	0.000	11:36
26 Std. Gas	0.080	11:36
27 Room Air	0.000	11:37
28 Std. Gas	0.080	11:37
29 Room Air	0.000	11:38
30 Std. Gas	0.080	11:38
31 Room Air	0.000	11:38
32 Std. Gas	0.080	11:39
33 Room Air	0.000	11:39
34 Std. Gas	0.080	11:40
35 Room Air	0.000	11:40
36 Std. Gas	0.080	11:41
37 Room Air	0.000	11:41
38 Std. Gas	0.080	11:41
39 Room Air	0.000	11:42
40 Std. Gas	0.079	11:42
41 Room Air	0.000	11:43
42 Cal Check Stats:		
43 Average	0.0800	
44 Std Dev	0.0006	
45 Rel Std Dev	0.82%	

*Stability test
Deb Kashner
p.1/2*

Operator Signature
DEB KASHUR

Remarks:

Lot #: 14512080A2 Cyl #: 22 Exp. date 08/01/2014

Form 106-I8000

Stability test
Deb Kashur
p. 2/2