

INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

Serial No.: 80-003057 Instrument Location: GFSO

Reason for Install/Repair: _____ Install After Lab Repair/Inspection
 Other (Specify) New Instrument installed.

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer®.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: N/A

11/14/09
Date

J. Oldenberg 274.
Field Inspector's Signature

04 Dec. 09
Date

Deb Kashner
Reviewed By

80-003057
P. 1 of 4

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = GFSO 8164.13.00 06/09
11/14/2009 21:47

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659357
Standard Cyl #? 26
Standard Expiration? 09/03/2011
Oper No? 048915

Flow Cal. Date: 04/07/2008
Slope 718
Intercept -644065

IR Calibration Date: 06/05/2008
 3um 9um

0th Coef(*100): -22884 -16302
1st Coef(*100): 265055 140885
2nd Coef(*100): 151 572
H2O adj(mg/l*10k): 508 293

***** Printer Test End *****



Operator Signature
DOLF OLDENBURG

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = GFSO 8164.13.00 06/09
11/14/2009 21:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:49
02 Std. Gas	0.081	21:49
03 Room Air	0.000	21:50
04 Std. Gas	0.081	21:50
05 Room Air	0.000	21:51
06 Std. Gas	0.082	21:51
07 Room Air	0.000	21:52


Lot No = 659357

Cyl No = 26

Exp Date = 09/03/2011

County = 18

Oper No. = 048915

 274.

Operator Signature
DOLF OLDENBURG

Remarks:

Form 106-I8000

80-003057
p. 3 of 4 YAK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003057
Location = GFSO 8164.13.00 06/09
11/14/2009 21:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	21:53
02 Room Air	0.000	21:54

*Invalid Test
Inhibited - RFI

Lot No = 659357
Cyl No = 26
Exp Date = 09/03/2011
County = 18 Oper No. = 048915



Operator Signature
DOLF OLDENBURG

Remarks:

Form 106-I8000

80-003057
p. 4 of 4