

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003069  
Location = ~~TOXL WARD~~      8164.13.00 06/09  
11/14/2009      *DK 21 April*      11:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:14
02 Std. Gas	0.080	11:14
03 Room Air	0.000	11:15
04 Std. Gas	0.080	11:15
05 Room Air	0.000	11:16
06 Std. Gas	0.081	11:16
07 Room Air	0.000	11:16

Lot No = 659358

Cyl No = 32

Exp Date = 09/03/2011

County = 51

Oper No. = 049509



Operator Signature  
PATRICK C. HUDSON

Remarks:

Form 106-I8000

*Revised*  
*80-003069*  
*P. 105 of 105*  
*KRM*

*80-003069*  
*P. 2 of 2*  
*KRM*