



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Duane L. Marmon

Location: STAN

Intoxilyzer Serial No.: 80-003061

Lot Number: 659358

Cylinder Number: 43

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %

Test 2: 0. 082 %

Test 3: 0. 082 %

(ATTACH TEST RECORD)

[Signature]
 Chemical Test Operator Signature
 Deb Kasher

10/14/09
 Date Cylinder Changed
 04 Dec 09

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-003061

P. 1 of 2
 FH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003061
Location = STAN 8164.12.00 06/09
11/14/2009 23:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	23:07
02 Std. Gas	0.081	23:07
03 Room Air	0.000	23:08
04 Std. Gas	0.082	23:08
05 Room Air	0.000	23:09
06 Std. Gas	0.082	23:09
07 Room Air	0.000	23:10


Lot No = 659358

Cyl No = 43

Exp Date = 09/03/2011

County = 31

Oper No. = 100019



Operator Signature
DUANE L. MARMON

Remarks:

Form 106-I8000

80-003061
P. 2 of 2
AK