



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Patrick C Hudson

Location: Ward County Jail

Intoxilyzer Serial No.: 80-003069

Lot Number: 659358

Cylinder Number: 32

Expiration Date: 9/3/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 080 %

Test 2: 0. 080 %

Test 3: 0. 081 %

(ATTACH TEST RECORD)

Patrick C Hudson
 Chemical Test Operator Signature

11/14/2009
 Date Cylinder Changed

Deb Kashner
 Reviewed By

17 NOV 2009
 Date

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

[Type text]

80-003069
 P. 1 of 2 gsh

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003069
Location = TOXL 8164.13.00 06/09
11/14/2009 11:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:14
02 Std. Gas	0.080	11:14
03 Room Air	0.000	11:15
04 Std. Gas	0.080	11:15
05 Room Air	0.000	11:16
06 Std. Gas	0.081	11:16
07 Room Air	0.000	11:16


Lot No = 659358

Cyl No = 32

Exp Date = 09/03/2011

County = 51

Oper No. = 049509



Operator Signature
PATRICK C. HUDSON

Remarks:

Form 106-I8000

80-003069
p. 2 of 2
PH