



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Lt. Tracy L. Meidinger #258

Location: UNDP

Intoxilyzer Serial No.: 80-004206

Lot Number: 659358

Cylinder Number: 31

Expiration Date: 09-03-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082 %

Test 2: 0. 082 %

Test 3: 0. 082 %

(ATTACH TEST RECORD)

Lt. Tracy L. Meidinger #258  
 Chemical Test Operator Signature

11-14-2009  
 Date Cylinder Changed

Deb Kashur  
 Reviewed By

17 NOV 2009  
 Date

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

[Type text]

80-004206  
 P. Lot 2 FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004206  
Location = UNDP      8164.13.00 06/09  
11/14/2009      19:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	19:22
02 Std. Gas	0.082	19:22
03 Room Air	0.000	19:23
04 Std. Gas	0.082	19:23
05 Room Air	0.000	19:24
06 Std. Gas	0.082	19:24
07 Room Air	0.000	19:25

Lot No = 659358  
Cyl No = 31  
Exp Date = 09/03/2011  
County = 18      Oper No. = 119409

*Tracy L. Meidinger #258*

Operator Signature  
TRACY L. MEIDINGER

Remarks:

*AEA-Instal*

Form 106-18000

*80-004206  
f. 2af 2 8m*