



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Lori Flaten</u>		
Location: <u>Mantoloking PD</u>		
Intoxilyzer® Serial Number: <u>80-004939</u>		
Gas Lot Number: <u>659378</u>	Gas Cylinder Number: <u>40</u>	Expiration Date: <u>09-03-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.081</u> %		
Test 2: <u>0.081</u> %		
Test 3: <u>0.082</u> %		
(ATTACH TEST RECORD)		


Chemical Test Operator Signature

11-15-10
Date

Deb Kashur
Reviewed By

17 Dec 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004939
Location = MAND 8164.13.00 06/09
11/15/2010 16:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:16
02 Std. Gas	0.081	16:17
03 Room Air	0.000	16:17
04 Std. Gas	0.081	16:18
05 Room Air	0.000	16:18
06 Std. Gas	0.082	16:18
07 Room Air	0.000	16:19

Lot No = 659358
Cyl No = 40
Exp Date = 09/03/2011
County = 30 Oper No. = 068019



Operator Signature
LORI E. FLATEN

Remarks:

Form 106-I8000

80-004939
P. 46 of 46
JAN