

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-002666

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.) ✓
Sim SN DR 5144 Lot No. — Rep AC 0.019
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.) ✓
Sim SN DR 5187 Lot No. — Rep AC 0.285
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.) ✓
Sim SN DR 5114 Lot No. — Rep AC 0.100
- J. RFI check. Run CMS mode. Key radio on first room air. (Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode. (Attach test record.)

Lot No. 659358 Cyl No. 46 Exp. Date 03 Sept 2011

Test 1 <u>0.081</u>	Test 4 <u>0.082</u>	Test 7 <u>0.082</u>
Test 2 <u>0.082</u>	Test 5 <u>0.082</u>	Test 8 <u>0.083</u>
Test 3 <u>0.082</u>	Test 6 <u>0.083</u>	Test 9 <u>0.082</u>

Average 0.082

L. Remarks/Maintenance record: Software ver 8164.13, annual inspection

16 NOV 2009
Date

Deb Keshur
Field Inspector's Signature

1 Dec 2009
Date

Mary Pearson
Reviewed by

P. 1 of 9
1/2/11

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
11/13/2009 11:20

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:21
02 Std. Sol.	0.019	11:22
03 Room Air	0.000	11:22
04 Std. Sol.	0.019	11:23
05 Room Air	0.000	11:23
06 Std. Sol.	0.020	11:24
07 Room Air	0.000	11:24

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC

Form 106-18000

P. 3 of 9
KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
11/13/2009 12:56

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:57
02 Std. Sol.	0.279	12:58
03 Room Air	0.000	12:58
04 Std. Sol.	0.281	12:59
05 Room Air	0.000	13:00
06 Std. Sol.	0.282	13:00
07 Room Air	0.000	13:01

08 Sim Temp = 34.0°C

Simul Ser No = ^{DR5187}~~00000~~ DK
Std Sol No = ~~000~~ 0.285
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Linearity test

Form 106-I8000

P. 4/2/9 SAM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
11/13/2009 13:27

Test	AC	Time
01 Room Air	0.000	13:27
02 *Subject Test	INT*	13:28
03 Room Air	0.000	13:28

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 46
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent test

Form 106-I8000

P. Safir

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
11/13/2009 16:34

Test	AC	Time
01 Diagnostic	OK	16:35
02 Room Air	RFI*	16:35
03 Room Air	0.000	16:36

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 46
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI Test

Form 106-18000

*P. Leaf 9
VLM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
11/16/2009 09:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:25
02 Std. Gas	0.081	09:26
03 Room Air	0.000	09:26
04 Std. Gas	0.082	09:27
05 Room Air	0.000	09:27
06 Std. Gas	0.082	09:27
07 Room Air	0.000	09:28

Lot No = 659358
Cyl No = 46
Exp Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-I8000

P. 7 of 9
KAM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
11/16/2009 09:29

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:30
02 Std. Gas	0.082	09:30
03 Room Air	0.000	09:31
04 Std. Gas	0.082	09:31
05 Room Air	0.000	09:32
06 Std. Gas	0.083	09:32
07 Room Air	0.000	09:33

Lot No = 659358
Cyl No = 46
Exp Date = 09/03/2011
County = 08 Oper No. = 777777



Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-I8000

*P. 8 of 9
sm*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-002666
Location = TOXL 8164.13.00 06/09
11/16/2009 09:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:38
02 Std. Gas	0.082	09:38
03 Room Air	0.000	09:39
04 Std. Gas	0.083	09:39
05 Room Air	0.000	09:40
06 Std. Gas	0.082	09:40
07 Room Air	0.000	09:40

Lot No = 659358
Cyl No = 46
Exp Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Calibration check
Form 106-I8000

*P. 9/2/9
12m*