

11/19/2009 10:11 NDOAG - Crime Laboratory Div

(FAX)701 328 6185

P.002/003



OFFICE OF ATTORNEY GENERAL  
Crime Laboratory Division  
2641 East Main Avenue  
P.O. Box 937  
Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Mark W. McDonald

Location: Valley City Police Dept.

Intoxilyzer Serial No.: 80-003058

Lot Number: 659358

Cylinder Number: 47

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.081 %

Test 2: 0.081 %

Test 3: 0.081 %

(ATTACH TEST RECORD)

[Signature]  
Chemical Test Operator Signature

11/17/09  
Date Cylinder Changed DK

[Signature]  
Reviewed By

24 NOV 2009  
Date

Form 105-G  
Toxicology Laboratory  
Crime Laboratory Division

[Type text]

80-003058  
p. 1 of 2  
RM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003058  
Location = VALL      8164.12.00 06/09  
11/17/2009      14:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:03
02 Std. Gas	0.081	14:03
03 Room Air	0.000	14:04
04 Std. Gas	0.081	14:04
05 Room Air	0.000	14:05
06 Std. Gas	0.081	14:05
07 Room Air	0.000	14:05

Lot No = 659358  
Cyl No = 47  
Exp Date = 09/03/2011  
County = 02      Oper No. = 119502



Operator Signature  
MARK W. MCDONALD

Remarks:

Form 106-18000

80-003058  
P. 2 of 2 FILE