



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David Feland</u>		
Location: <u>Kidder S.O.</u>		
Intoxilyzer® Serial Number: <u>80-004186 DK</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>37</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.081</u> %		
Test 2: <u>0.088</u> %		
Test 3: <u>0.082</u> %		
(ATTACH TEST RECORD)		

David Feland
Chemical Test Operator Signature

17 Nov 10
Date

Deb Kashur
Reviewed By

19 NOV 10 DK
Date

80-004186
P. 93-194
JAN


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
11/17/2010 14:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:49
02 Std. Gas	0.081	14:49
03 Room Air	0.000	14:50
04 Std. Gas	0.081	14:50
05 Room Air	0.000	14:50
06 Std. Gas	0.082	14:51
07 Room Air	0.000	14:51

Lot No = 659357
Cyl No = 37
Exp Date = 09/03/2011
County = 22 Oper No. = 119404



Operator Signature
DAVID FELAND

Remarks:

Form 106-I8000

80-004186
P. 94494
JFM