



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David M Wallace

Location: Dickinson, ND

Intoxilyzer Serial No.: 80-003066

Lot Number: 659358

Cylinder Number: 23

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082 %

Test 2: 0. 082 %

Test 3: 0. 082 %

(ATTACH TEST RECORD)

DM Wallace
 Chemical Test Operator Signature
 Deb Kashur

11-18-2009
 Date Cylinder Changed
 24 NOV 2009

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-003066
 P. 1 of 2
 Fh


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = DKPD 8164.13.00 06/09
11/18/2009 21:33

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:34
02 Std. Gas	0.082	21:34
03 Room Air	0.000	21:35
04 Std. Gas	0.082	21:35
05 Room Air	0.000	21:36
06 Std. Gas	0.082	21:36
07 Room Air	0.000	21:37

Lot No = 659358
Cyl No = 23
Exp Date = 09/03/2011
County = 45 Oper No. = 068332


Operator Signature
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-003066
P. 2 of 2 JCH