

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Richard D. Johnson

Location: McLean S.O. Washburn

Intoxilyzer Serial No.: 80-004196

Lot Number: 659357

Cylinder Number: 47

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.082 %

Test 2: 0.082 %

Test 3: 0.082 %

(ATTACH TEST RECORD)

Richard D. Johnson
Chemical Test Operator Signature

11/18/2009
Date

Deb Kashner
Reviewed By

15 Dec 2009
Date

80-004196
P.1 of 2 KM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004197
Location = WASH 8164.12.00 06/09
11/18/2009 01:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	01:21
02 Std. Gas	0.082	01:21
03 Room Air	0.000	01:21
04 Std. Gas	0.082	01:22
05 Room Air	0.000	01:22
06 Std. Gas	0.082	01:23
07 Room Air	0.000	01:23

Lot No = 659357
Cyl No = 47
Exp Date = 09/03/2011
County = 28 Oper No. = 118707


Operator Signature
RICHARD D. JOHNSON

Remarks:

Form 106-I8000

80-004196
P. 2 of 2 FR