

12/07/2009 13:28 NDOAG - Crime Laboratory Div

(FAX)701 328 6185

P.002/003

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

### ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Shannon Liebald

Location: Mercer County Sheriff's Dept.

Intoxilyzer Serial No.: 80-004194

Lot Number: 659357

Cylinder Number: 43

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082 %

Test 2: 0. 082 %

Test 3: 0. 081 %

(ATTACH TEST RECORD)

Shannon Liebald  
Chemical Test Operator Signature

11-19-09  
Date

Deb Kashner (ACA)  
Reviewed By

08 Dec 2009  
Date

80-004194  
P-1 of 2

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004194  
Location = STAT      8164.12.00 06/09  
11/19/2009      17:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:09
02 Std. Gas	0.082	17:09
03 Room Air	0.000	17:10
04 Std. Gas	0.082	17:10
05 Room Air	0.000	17:11
06 Std. Gas	0.081	17:11
07 Room Air	0.000	17:12


Lot No = 659357

Cyl No = 43

Exp Date = 09/03/2011

County = 29

Oper No. = 020904



Operator Signature  
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-604194  
p. 2 of 2 FM