

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: ^{DK} ~~68~~ 80-004184

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.) ✓
Sim SN DR5144 Lot No. Rep AC 0.019
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.) ✓
Sim SN DR5187 Lot No. Rep AC 0.285
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode. (Attach test record.) ✓
Sim SN DR5114 Lot No. Rep AC 0.100
- J. RFI check. Run CMS mode. Key radio on first room air. (Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode. (Attach test record.)

Lot No. 659358 Cyl No. 35 Exp. Date 03 Sept 2011

Test 1 <u>0.082</u>	Test 4 <u>0.081</u>	Test 7 <u>0.081</u>
Test 2 <u>0.081</u>	Test 5 <u>0.081</u>	Test 8 <u>0.082</u>
Test 3 <u>0.081</u>	Test 6 <u>0.081</u>	Test 9 <u>0.081</u>

Average 0.081

L. Remarks/Maintenance record: Software 8164.13 upgrade.

01 Dec 2009
Date
1 Dec 2009
Date

Deb Kasher
Field Inspector's Signature
Manny Plauer
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.13.00 06/09
11/20/2009 08:47

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659358
Standard Cyl #? 35
Standard Expiration? 09/03/2011
Oper No? 777777

Flow Cal. Date: 06/03/2009
Slope 696
Intercept -623534

IR Calibration Date: 06/03/2009
 3um 9um

0th Coef(*100): -10976 -21095
1st Coef(*100): 262845 132771
2nd Coef(*100): 1765 916
H2O adj(mg/l*10k): 544 397

***** Printer Test End *****

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Print test

Form 106-I8000

80-004184
P.2 of 9 FN

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.13.00 06/09
12/01/2009 08:49

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:49
02 Std. Sol.	0.017	08:50
03 Room Air	0.000	08:51
04 Std. Sol.	0.018	08:52
05 Room Air	0.000	08:52
06 Std. Sol.	0.018	08:53
07 Room Air	0.000	08:53

08 Sim Temp = 34.0°C

Simul Ser No = 00000
Std Sol No = 000
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Low AC

Form 106-I8000

*80-004184
p. 3 of 9 RW*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.13.00 06/09
12/01/2009 09:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:31
02 Std. Sol.	0.284	09:34
03 Room Air	0.000	09:34
04 Std. Sol.	0.290	09:35
05 Room Air	0.000	09:36
06 Std. Sol.	0.288	09:37
07 Room Air	0.000	09:37

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Linearity test 0.285 AC

Form 106-I8000

80-004184
P. 4 of 9

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.13.00 06/09
12/01/2009 10:25

Test	AC	Time
01 Room Air	0.000	10:25
02 *Subject Test	INT*	10:26
03 Room Air	0.000	10:27

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = 00000
Std Sol No = 000
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Interferent check

Form 106-I8000

80-004184
P. 549 FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.13.00 06/09
12/01/2009 10:43

Test	AC	Time
01 Diagnostic	OK	10:43
02 Room Air	RFI*	10:44
03 Room Air	0.000	10:44

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = 00000
Std Sol No = 000
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-I8000

*80-004184
p. 6 of 9 FM*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.13.00 06/09
12/01/2009 12:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:56
02 Std. Gas	0.082	12:57
03 Room Air	0.000	12:57
04 Std. Gas	0.081	12:58
05 Room Air	0.000	12:58
06 Std. Gas	0.081	12:59
07 Room Air	0.000	12:59

Lot No = 659358
Cyl No = 35
Exp Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-I8000

80-004184
P. 20/9 1/11

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.13.00 06/09
12/01/2009 13:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:03
02 Std. Gas	0.081	13:03
03 Room Air	0.000	13:04
04 Std. Gas	0.081	13:04
05 Room Air	0.000	13:05
06 Std. Gas	0.081	13:05
07 Room Air	0.000	13:06

Lot No = 659358
Cyl No = 35
Exp Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-18000

80-004184
P. 8 of 9 AK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = TOXL 8164.13.00 06/09
12/01/2009 13:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:08
02 Std. Gas	0.081	13:08
03 Room Air	0.000	13:09
04 Std. Gas	0.082	13:09
05 Room Air	0.000	13:10
06 Std. Gas	0.081	13:10
07 Room Air	0.000	13:10

Lot No = 659358
Cyl No = 35
Exp Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-I8000