



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (5-2011)

Chemical Test Operator Name (Print) <i>Melby, Robert M.</i>		
Location <i>CROS Divide Co. S.O.</i>	Intoxilyzer® Serial Number <i>80-004200</i>	
Gas Lot Number <i>23411080A1</i>	Gas Cylinder Number <i>13</i>	Gas Expiration Date <i>10/01/2013</i>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

- A. *0.079* AC
- B. *0.080* AC
- C. *0.079* AC

- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Breath Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>Robert M. Melby</i>	Date <i>11/23/11</i>
Reviewed By (Crime Laboratory Use Only) <i>Deb Kishner</i>	Date <i>03 Apr 12</i>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004200
Location = CROS 8164.13.00 06/09
11/23/2011 12:20

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:20
02 Std. Gas	0.079	12:21
03 Room Air	0.000	12:21
04 Std. Gas	0.080	12:22
05 Room Air	0.000	12:22
06 Std. Gas	0.079	12:22
07 Room Air	0.000	12:23

Lot No = 23411080A1
Cyl No = 13
Exp Date = 10/01/2013
County = 12 Oper No. = 100011



Operator Signature
ROBERT M. MELBY

Remarks:

Test OK Form 106-I8000