

INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

Serial No.: 80-004196 Instrument Location: Wash

Reason for Install/Repair: Install After Lab Repair/Inspection
 Other (Specify) _____

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer®.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____

11-24-09
Date

Richard Johnson
Field Inspector's Signature

04 Dec 09
Date

Deb Kashur
Reviewed By

80-004196
P. K. Y. M.

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = WASH 8164.13.00 06/09
11/24/2009 14:15

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659357
Standard Cyl #? 47
Standard Expiration? 09/03/2011
Oper No? 118707

Flow Cal. Date: 06/04/2009
Slope 727
Intercept -600779

IR Calibration Date: 06/03/2009
 3um 9um

0th Coef(*100):	-16078	-19878
1st Coef(*100):	252494	128485
2nd Coef(*100):	3196	1141
H2O adj(mg/l*10k):	686	431

***** Printer Test End *****



Operator Signature
RICHARD D. JOHNSON

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = WASH 8164.13.00 06/09
11/24/2009 14:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:17
02 Std. Gas	0.081	14:17
03 Room Air	0.000	14:18
04 Std. Gas	0.081	14:18
05 Room Air	0.000	14:19
06 Std. Gas	0.082	14:19
07 Room Air	0.000	14:19


Lot No = 659357

Cyl No = 47

Exp Date = 09/03/2011

County = 28

Oper No. = 118707


Operator Signature
RICHARD D. JOHNSON

Remarks:

Form 106-I8000

80-004196
p. 3 of 4 JRM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004196
Location = WASH 8164.13.00 06/09
11/24/2009 14:24


Test	AC	Time
01 Diagnostic	OK	14:26
02 Room Air	RFI*	14:26
03 Room Air	0.000	14:27

*Invalid Test
Inhibited - RFI

Sub Name = TEST, SUBJECT NA
Sub DOB = 07/03/1962
Sub Sex = Male Weight = 180
Test = OTH Cit = 11
Dr. Lic. = ND/11
Lot No = 659357
Cyl No = 47

Expiration Date = 09/03/2011
County = 28 Oper No. = 118707

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
RICHARD D. JOHNSON

Remarks:

Form 106-I8000

80-004196
P. 4 of 4 JH