



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>DUISEN, SHANNON</u>		
Location: <u>TIOGA</u>		
Intoxilyzer® Serial Number: <u>80-003071</u>		
Gas Lot Number: <u>15410080A</u>	Gas Cylinder Number: <u>25</u>	Expiration Date: <u>07/01/12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: <u>0.079</u> %		
Test 2: <u>0.078</u> %		
Test 3: <u>0.079</u> %		
(ATTACH TEST RECORD)		

[Signature]  
Chemical Test Operator Signature

11/24/10  
Date

Deb Kashur  
Reviewed By

30 NOV 2010  
Date

80-003071  
P. 106 of 107  
PM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003071  
Location = TIOG      8164.13.00 06/09  
11/24/2010      22:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	22:31
02 Std. Gas	0.079	22:31
03 Room Air	0.000	22:32
04 Std. Gas	0.078	22:32
05 Room Air	0.000	22:33
06 Std. Gas	0.079	22:33
07 Room Air	0.000	22:34

Lot No = 15410080A1  
Cyl No = 25  
Exp Date = 07/01/2012  
County = 53      Oper No. = 020632



Operator Signature  
SHANNON P. DUISEN

Remarks:

Form 106-I8000

80-003071  
P. 107 of 107  
SM