



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Darcy W Aberle

Location: Williston LEC

Intoxilyzer Serial No.: 80-002666

Lot Number: 659357

Cylinder Number: 45

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.081%

Test 2: 0.082%

Test 3: 0.080%

(ATTACH TEST RECORD)

Darcy W Aberle  
 Chemical Test Operator Signature

11-18-09  
 Date Cylinder Changed

Deb Kashner  
 Reviewed By

24 NOV 2009  
 Date

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

[Type text]

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-002666  
Location = WILL      8164.13.00 06/09  
11/18/2009      16:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:27
02 Std. Gas	0.081	16:27
03 Room Air	0.000	16:28
04 Std. Gas	0.082	16:28
05 Room Air	0.000	16:29
06 Std. Gas	0.081	16:29
07 Room Air	0.000	16:29

Lot No = 659357  
Cyl No = 45  
Exp Date = 09/03/2011  
County = 53      Oper No. = 010101

Darcy W Aberle  
Operator Signature  
DARCY W. ABERLE

Remarks:

ACA Test

Form 106-I8000