



**ETHANOL BREATH STANDARD CYLINDER REPORT**  
 OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>William R Petty</u>		
Location: <u>GFAF</u>		
Intoxilyzer® Serial Number: <u>80-002667</u>		
Gas Lot Number: <u>DK 65 659358</u>	Gas Cylinder Number: <u>39</u>	Expiration Date: <u>09/03/2011</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>081</u> % Test 2: 0. <u>080</u> % Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Will Petty  
 Chemical Test Operator Signature

25 Nov 10  
 Date

Deb Kashner  
 Reviewed By

22 Mar 2011  
 Date

80-002667  
 P. 82 of 83  
 FN


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-002667  
Location = GFAP                    8164.13.00 06/09  
11/25/2010                            10:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:01
02 Std. Gas	0.081	10:02
03 Room Air	0.000	10:02
04 Std. Gas	0.080	10:03
05 Room Air	0.000	10:03
06 Std. Gas	0.080	10:03
07 Room Air	0.000	10:04

Lot No = 659358  
Cyl No = 39  
Exp Date = 09/03/2011  
County = 18                            Oper No. = 120011

  
\_\_\_\_\_  
Operator Signature  
WILLIAM R. PETTY

Remarks: *ACA Test*

Form 106-I8000

*80-002667  
P. 83 of 83  
JRM*