



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Richard A Sherlock (4713)</u>		
Location: <u>WALSH COUNTY LEL GRAFTON ND 58237</u>		
Intoxilyzer® Serial Number: <u>80-003072</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>16</u>	Expiration Date: <u>09/03/11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>081</u> % Test 2: 0. <u>081</u> % Test 3: 0. <u>082</u> %		

(ATTACH TEST RECORD)

Richard A Sherlock (4713)
Chemical Test Operator Signature

11/25/10
Date

Deb Keshner
Reviewed By

06 Dec 2010
Date

80-003072
P. 98 of 99
JK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = GRAF 8164.13.00 06/09
11/25/2010 02:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	02:07
02 Std. Gas	0.081	02:07
03 Room Air	0.000	02:08
04 Std. Gas	0.081	02:08
05 Room Air	0.000	02:09
06 Std. Gas	0.082	02:09
07 Room Air	0.000	02:10

Lot No = 659358
Cyl No = 16
Exp Date = 09/03/2011
County = 50 Oper No. = 100428

 (4713)

Operator Signature
RICHARD A. SHERLOCK

Remarks: *ALA - PASS*

Form 106-I8000

80-003072
P. 99 of 99
KAM