



ETHANOL BREATH STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Lt. Tracy L. Meidinger #258</u>		
Location: <u>UND-PD</u>		
Intoxilyzer® Serial Number: <u>80-004934</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>31</u>	Expiration Date: <u>09-30-11</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: 0. <u>081</u> % Test 2: 0. <u>081</u> % Test 3: 0. <u>081</u> %		
(ATTACH TEST RECORD)		

Lt. Tracy L. Meidinger #258
 Chemical Test Operator Signature

11-25-10
 Date

Deb Kashner
 Reviewed By

30 NOV 2010
 Date

80-004934
 P. Col of 62
 sm

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004934
Location = UNDP 8164.13.00 06/09
11/25/2010 18:10

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:11
02 Std. Gas	0.081	18:11
03 Room Air	0.000	18:11
04 Std. Gas	0.081	18:12
05 Room Air	0.000	18:12
06 Std. Gas	0.081	18:13
07 Room Air	0.000	18:13

Lot No = 659358
Cyl No = 31
Exp Date = 09/03/2011
County = 18 Oper No. = 119409

Tracy L. Meidinger #258
Operator Signature
TRACY L. MEIDINGER

Remarks:

ACA 11-25-10

Form 106-I8000

80-004934
P. 62 of 62
VFM