

12/07/2009 15:13 NDOAG - Crime Laboratory Div

(FAX)701 328 6185

P.002/004

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Shannon Liebold

Location: Mercer County Sheriff's Dept.

Intoxilyzer Serial No.: 80-004194

Lot Number: 659357

Cylinder Number: 43

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 079 %

Test 2: 0. 081 %

Test 3: 0. 080 %

(ATTACH TEST RECORD)

Shannon Liebold  
Chemical Test Operator Signature

11-29-09  
Date

Deb Kashur (AEA)  
Reviewed By

18 Dec 2009  
Date

80-004194  
P. 1 of 2 KAL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004194  
Location = STAT      8164.12.00 06/09  
11/29/2009      00:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	00:53
02 Std. Gas	0.079	00:53
03 Room Air	0.000	00:54
04 Std. Gas	0.081	00:54
05 Room Air	0.000	00:55
06 Std. Gas	0.080	00:55
07 Room Air	0.000	00:55

Lot No = 659357

Cyl No = 43

Exp Date = 09/03/2011

County = 29

Oper No. = 020904



Operator Signature  
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-004194  
P. 2 of 2 FM