



# ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION  
Toxicology Section/Breath Alcohol Program  
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>David M Walker</u>		
Location: <u>Dickinson ZPC</u>		
Intoxilyzer® Serial Number: <u>80-004947</u>		
Gas Lot Number: <u>154108021</u>	Gas Cylinder Number: <u>20</u>	Expiration Date: <u>7-1-12</u>
<input type="checkbox"/> New Cylinder Installed <input checked="" type="checkbox"/> Cylinder Not Changed		
Instrument Test Results (Report to 3 Digits; Example, 0.081%) Test 1: <del>0.075%</del> <u>0.079%</u> <sup>DK</sup> Test 2: <u>0.079%</u> Test 3: <u>0.079%</u>		
(ATTACH TEST RECORD)		

David M Walker  
Chemical Test Operator Signature

12-1-10  
Date

Deb Kashur  
Reviewed By

06 Dec 2010  
Date

80-004947  
P. 55 of 56  
JCM


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004947  
Location = DKPD      8164.13.00 06/09  
12/01/2010      07:13

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:13
02 Std. Gas	0.079	07:14
03 Room Air	0.000	07:14
04 Std. Gas	0.079	07:15
05 Room Air	0.000	07:15
06 Std. Gas	0.079	07:15
07 Room Air	0.000	07:16

Lot No = 15410080A1  
Cyl No = 20  
Exp Date = 07/01/2012  
County = 45      Oper No. = 068332



Operator Signature  
DAVID M. WALLACE

Remarks:

Form 106-I8000

80-004947  
P. 56 of 56  
JEM