

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-003072

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)
Sim SN DR5144 Lot No. — Rep AC 0.019 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)
Sim SN DR5187 Lot No. — Rep AC 0.285 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.
(Attach test record.)
Sim SN DR5114 Lot No. — Rep AC 0.100 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.
(Attach test record.)

Lot No. 659357 Cyl No. 11 Exp. Date 03 Sept 2011

Test 1 0.082 Test 4 0.082 Test 7 0.083

Test 2 0.082 Test 5 0.081 Test 8 0.081

Test 3 0.083 Test 6 0.082 Test 9 0.082

Average 0.082

L. Remarks/Maintenance record: Software 8164.13 installed. Annual inspection

02 Dec 2009
Date

Deb Keshur
Field Inspector's Signature

4 Dec 2009
Date

Mary Pearson
Reviewed by

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
12/01/2009 08:28

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659357
Standard Cyl #? 11
Standard Expiration? 09/03/2011
Oper No? 777777

Flow Cal. Date: 06/05/2008
Slope 732
Intercept -720104

IR Calibration Date: 06/05/2008

	3um	9um
0th Coef(*100):	-20044	-17692
1st Coef(*100):	264480	139417
2nd Coef(*100):	1329	776
H2O adj(mg/l*10k):	540	318

***** Printer Test End *****

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Print test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
12/01/2009 14:01

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:01
02 Std. Sol.	0.019	14:02
03 Room Air	0.000	14:03
04 Std. Sol.	0.018	14:03
05 Room Air	0.000	14:04
06 Std. Sol.	0.018	14:05
07 Room Air	0.000	14:05

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Low AC

Form 106-18000

*80-003072
P. 349 102*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
12/02/2009 07:57

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	07:57
02 Std. Sol.	0.283	07:59
03 Room Air	0.000	08:00
04 Std. Sol.	0.286	08:00
05 Room Air	0.000	08:01
06 Std. Sol.	0.285	08:02
07 Room Air	0.000	08:02

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher
Operator Signature
DEB KASHUR

Remarks:

Linearity

Form 106-I8000

80-003072
P. 4 of 9 JAK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
12/02/2009 08:04

Test	AC	Time
01 Room Air	0.000	08:05
02 *Subject Test	INT*	08:05
03 Room Air	0.000	08:06

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Simul Ser No = 00000
Std Sol No = 000
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Interferent test

Form 106-I8000

80-003072
P. 5/7 YAM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
12/02/2009 08:25

Test	AC	Time
01 Diagnostic	OK	08:26
02 Room Air	RFI*	08:26
03 Room Air	0.000	08:26

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659357
Cyl No = 11
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI Test

Form 106-I8000

80-003072
P. 6 of 9 JSA

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
12/01/2009 13:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:32
02 Std. Gas	0.082	13:32
03 Room Air	0.000	13:33
04 Std. Gas	0.082	13:33
05 Room Air	0.000	13:33
06 Std. Gas	0.083	13:34
07 Room Air	0.000	13:34

Lot No = 659357
Cyl No = 11
Exp Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-I8000

80-003072
A 749 KM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
12/01/2009 13:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:40
02 Std. Gas	0.082	13:41
03 Room Air	0.000	13:41
04 Std. Gas	0.081	13:41
05 Room Air	0.000	13:42
06 Std. Gas	0.082	13:42
07 Room Air	0.000	13:43

Lot No = 659357
Cyl No = 11
Exp Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:
Calibration check

Form 106-I8000

*80-003072
P. 8/9/10*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003072
Location = TOXL 8164.13.00 06/09
12/01/2009 13:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:44
02 Std. Gas	0.083	13:45
03 Room Air	0.000	13:45
04 Std. Gas	0.081	13:46
05 Room Air	0.000	13:46
06 Std. Gas	0.082	13:46
07 Room Air	0.000	13:47

Lot No = 659357

Cyl No = 11

Exp Date = 09/03/2011

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Calibration check

Form 106-I8000

80-003072
P. 949 JAM