



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Jacquie DeMarce</u>		
Location: <u>FTOT</u>		
Intoxilyzer® Serial Number: <u>80-004194</u>		
Gas Lot Number: <u>659358</u>	Gas Cylinder Number: <u>56</u>	Expiration Date: <u>9/03/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>077</u> %		
Test 2: 0. <u>078</u> %		
Test 3: 0. <u>077</u> %		
(ATTACH TEST RECORD)		

Jacquie DeMarce

Chemical Test Operator Signature

12-2-10

Date

Deb Kashner

Reviewed By

06 Dec 2010

Date

80-004194
p. 87 of 88
RM


Intoxilyzer Test Record and Checklist
NDDAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = FTOT 8164.13.00 06/09
12/02/2010 08:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:19
02 Std. Gas	0.077	08:19
03 Room Air	0.000	08:19
04 Std. Gas	0.076	08:19
05 Room Air	0.000	08:20
06 Std. Gas	0.077	08:20
07 Room Air	0.000	08:20

Lot No = 659358
Cyl No = 56
Exp Date = 09/03/2011
County = 03 Oper No. = 100103


Operator Signature
JACQUIE DEMARCE

Remarks:

Form 106-18000

80-00 4194
P. 88 of 88
AM