

INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

Serial No.: 80-004184 Instrument Location: Valley City Police Dept.

Reason for Install/Repair: _____ Install After Lab Repair/Inspection
X Other (Specify) new instrument

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer®.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____

12/03/2009
Date

[Signature]
Field Inspector's Signature

04 Dec 2009
Date

Deb Kashner
Reviewed By

80-004184
P. 1 of 4 AL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = VALL 8164.13.00 06/09
12/03/2009 09:17

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659358
Standard Cyl #? 47
Standard Expiration? 09/03/2011
Oper No? 119502

Flow Cal. Date: 06/03/2009
Slope 696
Intercept -623534

IR Calibration Date: 06/03/2009

	3um	9um
0th Coef(*100):	-10976	-21095
1st Coef(*100):	262845	132771
2nd Coef(*100):	1765	916
H2O adj(mg/l*10k):	544	397

***** Printer Test End *****



Operator Signature
MARK W. MCDONALD

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = VALL 8164.13.00 06/09
12/03/2009 09:18

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:19
02 Std. Gas	0.079	09:19
03 Room Air	0.000	09:20
04 Std. Gas	0.081	09:20
05 Room Air	0.000	09:21
06 Std. Gas	0.081	09:21
07 Room Air	0.000	09:22

Lot No = 659358
Cyl No = 47
Exp Date = 09/03/2011
County = 02 Oper No. = 119502



Operator Signature
MARK W. MCDONALD

Remarks:

Form 106-I8000

80-004184
p. 3 of 4 JH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = VALL 8164.13.00 06/09
12/03/2009 09:23

Test	AC	Time
01 Diagnostic	OK	09:25
02 Room Air	RFI*	09:25
03 Room Air	0.000	09:25

*Invalid Test
Inhibited - RFI

Sub Name = TEST, SUBJECT NA
Sub DOB = 12/12/1912
Sub Sex = Male Weight = 123
Test = OTH Cit = 123
Dr. Lic. = ND/123
Lot No = 659358
Cyl No = 47
Expiration Date = 09/03/2011
County = 02 Oper No. = 119502

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
MARK W. MCDONALD

Remarks:

Form 106-I8000

80-004184
P. 4 of 4 AM