

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Mark W. McDonald

Location: Delley City Police Dept.

Intoxilyzer Serial No.: 80-004184

Lot Number: 0593 58

Cylinder Number: 47

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.061 %

Test 2: 0.061 %

Test 3: 0.062 %

(ATTACH TEST RECORD)


Chemical Test Operator Signature

12/03/2011
Date

Deb Keshu
Reviewed By

04 Dec 2009
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = VALL 8164.13.00 06/09
12/03/2009 09:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:27
02 Std. Gas	0.081	09:27
03 Room Air	0.000	09:28
04 Std. Gas	0.081	09:29
05 Room Air	0.000	09:29
06 Std. Gas	0.082	09:30
07 Room Air	0.000	09:30

Lot No = 659358
Cyl No = 47
Exp Date = 09/03/2011
County = 02 Oper No. = 119502

Operator Signature
MARK W. MCDONALD

Remarks:

Form 106-I8000

80-004184
P. 2 of 2

12/17/2009 15:04 NDOAG - Crime Laboratory Div

(FAX)701 328 6185

P.002/002

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004184
Location = VALL 8164.13.00 06/09
12/03/2009 09:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:27
02 Std. Gas	0.081	09:27
03 Room Air	0.000	09:28
04 Std. Gas	0.081	09:29
05 Room Air	0.000	09:29
06 Std. Gas	0.082	09:30
07 Room Air	0.000	09:30

Lot No = 659358
Cyl No = 47
Exp Date = 09/03/2011
County = 02 Oper No. = 119502



Operator Signature
MARK W. MCDONALD

Remarks:

Form 106-18000

Added page 3
30DEC09 FM
80-604184
P. 2 of 2