



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Shannon Liebold</u>		
Location: <u>Mercer County Sheriff's Dept.</u>		
Intoxilyzer® Serial Number: <u>80-003060</u>		
Gas Lot Number: <u>659357</u>	Gas Cylinder Number: <u>43</u>	Expiration Date: <u>9-3-11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	-
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>081</u> %		
Test 2: 0. 081 % <u>0.082%</u> DK		
Test 3: 0. <u>080</u> %		
(ATTACH TEST RECORD)		

Shannon Liebold
Chemical Test Operator Signature

12-3-10
Date

Deb Kasheer
Reviewed By

10 Dec '10
Date

80-003060
P. 96 of 97
SKM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003060
Location = STAT 8164.13.00 06/09
12/03/2010 18:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:52
02 Std. Gas	0.081	18:52
03 Room Air	0.000	18:53
04 Std. Gas	0.082	18:53
05 Room Air	0.000	18:54
06 Std. Gas	0.080	18:54
07 Room Air	0.000	18:55

Lot No = 659357

Cyl No = 43

Exp Date = 09/03/2011

County = 29

Oper No. = 020904



Operator Signature
SHANNON P. LIEBOLD

Remarks:

Form 106-I8000

80-003060
P. 97 of 97
JAN