



OFFICE OF ATTORNEY GENERAL  
 Crime Laboratory Division  
 2641 East Main Avenue  
 P.O. Box 937  
 Bismarck, North Dakota 58502-0937

TEL 701-328-6159

**ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Aaron R. Hummel

Location: GRAF

Intoxilyzer Serial No.: 80-003072

Lot Number: 659358

Cylinder Number: 16

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.082 %

Test 2: 0.082 %

(ATTACH TEST RECORD)

Test 3: 0.083 %

[Signature]  
 Chemical Test Operator Signature

12/4/09  
 Date Cylinder Changed

Deb Kashner  
 Reviewed By

15 Dec 09  
 Date

Form 105-G  
 Toxicology Laboratory  
 Crime Laboratory Division

80-003072  
 P. 1 of 2 RL

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003072  
Location = GRAF      8164.13.00 06/09  
12/04/2009      14:22

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:22
02 Std. Gas	0.082	14:23
03 Room Air	0.000	14:23
04 Std. Gas	0.082	14:24
05 Room Air	0.000	14:24
06 Std. Gas	0.083	14:25
07 Room Air	0.000	14:25


Lot No = 659358

Cyl No = 16

Exp Date = 09/03/2011

County = 50

Oper No. = 049907



Operator Signature  
AARON HUMMEL

Remarks:

*Cylinder Install ACA*

Form 106-I8000

80-003072  
P. 2 of 2 KH