

INTOXILYZER 5000 G and 8000 INSPECTION

Serial No.: 80-004197

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)  
Sim SN DRS144 Lot No.      Rep AC 0.019 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)  
Sim SN DRS187 Lot No.      Rep AC 0.285 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.  
(Attach test record.)  
Sim SN DRS114 Lot No.      Rep AC 0.106 ✓
- J. RFI check. Run CMS mode. Key radio on first room air.  
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.)

Lot No. 659358 Cyl No. 54 Exp. Date 03 Sept 2011

Test 1 <u>0.083</u>	Test 4 <u>0.083</u>	Test 7 <u>0.083</u>
Test 2 <u>0.083</u>	Test 5 <u>0.084</u>	Test 8 <u>0.084</u>
Test 3 <u>0.084</u>	Test 6 <u>0.084</u>	Test 9 <u>0.083</u>

Average 0.083

L. Remarks/Maintenance record: Software upgrade 8164.13. Annual inspection

04 Dec 2009 Date Deb Kashner Field Inspector's Signature  
4 Dec 2009 Date Mary Pearson Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004197  
Location = TOXL      8164.13.00 06/09  
12/04/2009      08:58

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:59
02 Std. Sol.	0.018	08:59
03 Room Air	0.000	09:00
04 Std. Sol.	0.018	09:01
05 Room Air	0.000	09:01
06 Std. Sol.	0.018	09:02
07 Room Air	0.000	09:02

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

Low AC 0.019 AC

Form 106-I8000

80-004197  
P 3 of 9 JTH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004197  
Location = TOXL      8164.13.00 06/09  
12/04/2009      09:25

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:26
02 Std. Sol.	0.289	09:27
03 Room Air	0.000	09:27
04 Std. Sol.	0.290	09:28
05 Room Air	0.000	09:29
06 Std. Sol.	0.292	09:29
07 Room Air	0.000	09:30

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test      0.285 AC*

Form 106-18000

*80-004197  
P. 4 of 9*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004197  
Location = TOXL                    8164.13.00 06/09  
12/04/2009                            10:46

Test	AC	Time
01 Room Air	0.000	10:47
02 *Subject Test	INT*	10:47
03 Room Air	0.000	10:47

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 54  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Interferent test*

Form 106-I8000

80-004197  
D. Sep 9, 2011

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004197  
Location = TOXL                    8164.13.00 06/09  
12/04/2009                            14:00

Test	AC	Time
01 Diagnostic	OK	14:01
02 Room Air	RFI*	14:01
03 Room Air	0.000	14:02

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 54  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-I8000

80-004197  
P. 6 of 7 JK

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004197  
Location = TOXL      8164.13.00 06/09  
12/04/2009      11:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:07
02 Std. Gas	0.083	11:08
03 Room Air	0.000	11:08
04 Std. Gas	0.083	11:08
05 Room Air	0.000	11:09
06 Std. Gas	0.084	11:09
07 Room Air	0.000	11:10

Lot No = 659358  
Cyl No = 54  
Exp Date = 09/03/2011  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*calibration check*

Form 106-I8000

80004197  
P. 7/2/09

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004197  
Location = TOXL                    8164.13.00 06/09  
12/04/2009                            12:46

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:46
02 Std. Gas	0.083	12:47
03 Room Air	0.000	12:47
04 Std. Gas	0.084	12:47
05 Room Air	0.000	12:48
06 Std. Gas	0.084	12:48
07 Room Air	0.000	12:49

Lot No = 659358  
Cyl No = 54  
Exp Date = 09/03/2011  
County = 08                            Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check*

Form 106-I8000

80-004197  
*P. 1st 9/09*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004197  
Location = TOXL      8164.13.00 06/09  
12/04/2009      13:25

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:25
02 Std. Gas	0.083	13:25
03 Room Air	0.000	13:26
04 Std. Gas	0.084	13:26
05 Room Air	0.000	13:27
06 Std. Gas	0.083	13:27
07 Room Air	0.000	13:28

Lot No = 659358  
Cyl No = 54  
Exp Date = 09/03/2011  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:  
*calibration check*

Form 106-I8000

*80-004197  
P. 949 JAW*