



ETHANOL BREATH STANDARD CYLINDER REPORT

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program
SFN 59282 (1-2010)

Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.

Chemical Test Operator Name: <u>Tyler Wolfe</u>		
Location: <u>HARV</u>		
Intoxilyzer® Serial Number: <u>80-004935</u>		
Gas Lot Number: <u>709352</u>	Gas Cylinder Number: <u>31</u>	Expiration Date: <u>12/23/11</u>
<input type="checkbox"/> New Cylinder Installed	<input checked="" type="checkbox"/> Cylinder Not Changed	
Instrument Test Results (Report to 3 Digits; Example, 0.081%)		
Test 1: 0. <u>076</u> %		
Test 2: 0. <u>077</u> %		
Test 3: 0. <u>077</u> %		
(ATTACH TEST RECORD)		

Tyler Wolfe
Chemical Test Operator Signature

12/4/10
Date

Deb Kashur
Reviewed By

06 Dec 2010
Date

80-004935
P. 69 of 65
FM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004935
Location = HARV 8164.13.00 06/09
12/04/2010 10:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:36
02 Std. Gas	0.076	10:36
03 Room Air	0.000	10:37
04 Std. Gas	0.077	10:37
05 Room Air	0.000	10:38
06 Std. Gas	0.077	10:38
07 Room Air	0.000	10:38

Lot No = 709352
Cyl No = 31
Exp Date = 12/23/2011
County = 52 Oper No. = 060901



Operator Signature
TYLER A. WOLFE

Remarks:

Form 106-I8000

80-004935
A 65265
JTW