

**ETHANOL BREATH STANDARD CYLINDER REPORT**

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Norman Reed

Location: Mercer Co SO - Stanton

Intoxilyzer Serial No.: ~~659357~~ 80-003060 DK

Lot Number: 659357

Cylinder Number: 43

Expiration Date: 09/03/2011


Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.080 %

Test 2: 0.081 %

Test 3: 0.081 %

(ATTACH TEST RECORD)

  
Chemical Test Operator Signature

12/05/2009  
Date

Deb Kashur  
Reviewed By

07 Dec 09  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003060  
Location = STAT      8164.13.00 06/09  
12/05/2009      16:30

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:32
02 Std. Gas	0.080	16:32
03 Room Air	0.000	16:32
04 Std. Gas	0.081	16:33
05 Room Air	0.000	16:33
06 Std. Gas	0.081	16:34
07 Room Air	0.000	16:34


Lot No = 659357

Cyl No = 43

Exp Date = 09/03/2011

County = 29

Oper No. = 049121



Operator Signature  
NORMAN C. RUUD

Remarks:

Form 106-I8000

80-003060  
P. 2 of 2 SEM