

OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 Instrument Certification and Maintenance Unit

BrF-005-8

✓ Q. Flow Monitor. Level 3 Func D Sub F. R 143 F 0.000 L/s

✓ R. Temperature Monitor. Level 3 Func D Sub T
 C 47.0 A 48.0 I 0.0
 B 44.9 V 0.0

✓ S. Gas Tank Monitor. Level 3 Func D Sub G
✓ T. Configure Wet Bath Sim. Level 1 Func S
329 psi Intox 330 psi Reg

✓ U. Zero Alcohol Test Level 1 Func C
DR3409 Sim SN — Lot 0.000 Rep Value
0.000 AC 0.000 AC 0.000 AC

✓ V. Low AC Test. Level 1 Func C
DR5144 Sim SN — Lot 0.019 Rep Value
0.019 AC 0.020 AC 0.020 AC

✓ W. Linearity Test. Level 1 Func C
DR5187 Sim SN — Lot 0.285 Rep Value
0.285 AC 0.288 AC 0.289 AC

✓ X. Interferent Test.
 Level 1 Func B
 Acetone DR 3845 Sim SN — Lot 0.0590 Conc
✓ Display ✓ Printout
 Methanol G10573 Sim SN — Lot 0.1090 Conc
✓ Display ✓ Printout
 IPA DR 5191 Sim SN — Lot 0.1090 Conc
✓ Display ✓ Printout

✓ Y. Ambient Test
✓ Display ✓ Printout

✓ Z. Invalid Sample Test
✓ Display ✓ Printout

✓ AA. RFI Detector Test. CMS Mode
✓ Display ✓ Printout

✓ BB. Configure Dry Gas Standard. Level 1 Func S

✓ CC. Stability Test at 0.080AC. Level 3 Func T Sub S
0.0823 Ave. 0.0004 Std Dev 0.55% Rel Std Dev

✓ DD. Maintenance/Repairs Software 8164.13 upgrade

08 Dec 2009
 Date
8 Dec 2009
 Date

Deb Kasher
 Signature
Manny Pearson
 Reviewed by

08/09

80-003065
 P. 2417
 FH

Intoxilyzer Test Record and Checklist
ND00AG Crime Lab. Div., Bismarck, ND 58501

EMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/07/2009 16:46

***** Printer Test *****

12 char/line
abcdefghijkl
ABCDEFGHIJKL

16 char per line
abcdefghijklmnop
ABCDEFGHIJKLMN

21 chars per line
abcdefghijklmnopqrstu
ABCDEFGHIJKLMNQRSTU

24 characters per line
abcdefghijklmnopqrstuwx
ABCDEFGHIJKLMNQRSTUWX

32 characters per line
abcdefghijklmnopqrstuwx123456
ABCDEFGHIJKLMNQRSTUWXYZ123456

42 characters per line
abcdefghijklmnopqrstuwx1234567890abcdef
ABCDEFGHIJKLMNQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup
Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Reslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Manfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659358
Standard Cyl #? 35
Standard Expiration? 09/03/2011
Oper No? 777777

Flow Cal. Date: 04/08/2008
Slope 694
Intercept -539075

IR Calibration Date: 06/05/2008
Sun Sun


0th Coef(*100): -9383 -9235

1st Coef(*100): 288128 138520

2nd Coef(*100): 957 781

#20 adj(mg/l+10k): 391 239

***** Test End *****


Operator Signature
DEB KASHUR

Remarks:


Form 106-18000

80-003065
P. 3 of 17 PM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/08/2009 08:32

Test	AC	Time
01 Room Air	0.000	08:33
02 *Subject Test	REF*	08:33
03 Room Air	0.000	08:34

*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 35
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Keshur

Operator Signature
DEB KASHUR

Remarks:

camera / Refusal check

Form 106-I8000

80-003065
P. 54/17844

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/07/2009 16:55

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 777777

Deb Keshur

Operator Signature
DEB KASHUR

Remarks:

Diagnostic

Form 106-I8000

*80-003065
A 6 of 17 fm*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/08/2009 08:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:39
02 Std. Sol.	0.000	08:40
03 Room Air	0.000	08:40
04 Std. Sol.	0.000	08:41
05 Room Air	0.000	08:42
06 Std. Sol.	0.000	08:42
07 Room Air	0.000	08:43

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

Zero alcohol test

Form 106-18000

80-003065
P. 7 of 17 FM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/08/2009 08:45

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:46
02 Std. Sol.	0.019	08:49
03 Room Air	0.000	08:50
04 Std. Sol.	0.020	08:51
05 Room Air	0.000	08:51
06 Std. Sol.	0.020	08:52
07 Room Air	0.000	08:52

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Low AC 0.019 AC

Form 106-I8000

80-003065
P. 8 of 12 YAK

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/08/2009 08:59

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:00
02 Std. Sol.	0.285	09:01
03 Room Air	0.000	09:02
04 Std. Sol.	0.288	09:02
05 Room Air	0.000	09:03
06 Std. Sol.	0.289	09:04
07 Room Air	0.000	09:04

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kasher

Operator Signature
DEB KASHUR

Remarks:

Linearity test ^{OK} 0.6 0.285 AC

Form 106-18000

80-003065
P. 9 of 17 fm

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/08/2009 09:15

Test	AC	Time
01 Room Air	0.000	09:15
02 *Subject Test	INT*	09:16
03 Room Air	0.000	09:16

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 35
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

acetone interferent

Form 106-I8000

80-003065
P. 10-17 8/24

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/08/2009 09:26

Test	AC	Time
01 Room Air	0.000	09:27
02 *Subject Test	INT*	09:27
03 Room Air	0.000	09:28

*Invalid Test
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 35
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

IPA ^{DK} ~~the~~ interferent

Form 106-I8000

80-003065
P. 12/17/09

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/08/2009 09:31

Test	AC	Time
01 Room Air	AMB*	09:32
02 Room Air	0.000	09:33

*Invalid Test
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 35
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

Deb Kashur
Operator Signature
DEB KASHUR

Remarks:

Ambient test

Form 106-I8000

80-003065
P. 13 of 17 JH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/08/2009 09:39

Test	AC	Time
01 Diagnostic	OK	09:39
02 Room Air	RFI*	09:40
03 Room Air	0.000	09:40

*Invalid Test
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT
Sub DOB = 02/01/1992
Sub Sex = Female Weight = 150
Test = DUI Cit = NA
Dr. Lic. = ND/DIS921456
Lot No = 659358
Cyl No = 35
Expiration Date = 09/03/2011
County = 08 Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur

Operator Signature
DEB KASHUR

Remarks:

RFI test

Form 106-18000

*80-003065
P. 15 of 17 yka*

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = TOXL 8164.13.00 06/09
12/07/2009 16:57

Test	AC	Time
01 Room Air	0.000	16:58
02 Std. Gas	0.082	16:58
03 Room Air	0.000	16:59
04 Std. Gas	0.082	16:59
05 Room Air	0.000	16:59
06 Std. Gas	0.083	17:00
07 Room Air	0.000	17:00
08 Std. Gas	0.082	17:01
09 Room Air	0.000	17:01
10 Std. Gas	0.082	17:02
11 Room Air	0.000	17:02
12 Std. Gas	0.083	17:02
13 Room Air	0.000	17:03
14 Std. Gas	0.082	17:03
15 Room Air	0.000	17:04
16 Std. Gas	0.082	17:04
17 Room Air	0.000	17:05
18 Std. Gas	0.082	17:05
19 Room Air	0.000	17:06
20 Std. Gas	0.083	17:06
21 Room Air	0.000	17:07
22 Std. Gas	0.082	17:07
23 Room Air	0.000	17:07
24 Std. Gas	0.083	17:08
25 Room Air	0.000	17:08
26 Std. Gas	0.083	17:09
27 Room Air	0.000	17:09
28 Std. Gas	0.082	17:09
29 Room Air	0.000	17:10
30 Std. Gas	0.082	17:10
31 Room Air	0.000	17:11
32 Std. Gas	0.082	17:11
33 Room Air	0.000	17:12
34 Std. Gas	0.082	17:12
35 Room Air	0.000	17:13
36 Std. Gas	0.082	17:13
37 Room Air	0.000	17:14
38 Std. Gas	0.082	17:14
39 Room Air	0.000	17:14
40 Std. Gas	0.082	17:15
41 Room Air	0.000	17:15
42 Cal Check Stats:		
43 Average	0.0823	
44 Std Dev	0.0004	
45 Rel Std Dev	0.55%	

*Stability test
Deb Kashner
p. 1082*

*80-003065
p. 16 of 17
JKH*

Operator Signature
DEB KASHUR

Remarks:

Lot #: 659358 Cyl #: 35

Exp date: 03 Sept 2011

Form 106-I8000

Stability test
Debs Kashur
p. 2 of 2

80-003065
P. 17 of 17