



**INTOXILYZER 5000 G and 8000 INSPECTION**

Serial No.: 80-003065

Location: TOXL

Check When Done

- A. Is the warm-up time less than 20 minutes? ✓
- B. Is a three-pronged grounded outlet used? ✓
- C. Is the breath tube heated? ✓
- D. Is the diagnostic check complete? ✓
- E. Is the time, date, and year correct (re-set if necessary)? ✓
- F. Print test. (Attach test record.) ✓
- G. Low AC. Use <0.03 AC in ACA mode. (Attach test record.)  
Sim SN DR5144 Lot No.      Rep AC 0.019 ✓
- H. Linearity Test. Use ≥ 0.25 AC in ACA mode. (Attach test record.)  
Sim SN DR5187 Lot No.      Rep AC 0.285 ✓
- I. Interferent check. Use 0.05% acetone plus 0.10 AC ethanol in ABA mode.  
(Attach test record.)  
Sim SN DR5114 Lot No.      Rep AC      ✓
- J. RFI check. Run CMS mode. Key radio on first room air.  
(Attach test record.) ✓
- K. Calibration Check. Use Ethanol Breath Standard cylinder. Do three sets in ACA mode.  
(Attach test record.)

Lot No. 659358 Cyl No. 35 Exp. Date 03 Sept 2011

Test 1 <u>0.083</u>	Test 4 <u>0.083</u>	Test 7 <u>0.083</u>
Test 2 <u>0.082</u>	Test 5 <u>0.083</u>	Test 8 <u>0.082</u>
Test 3 <u>0.082</u>	Test 6 <u>0.083</u>	Test 9 <u>0.083</u>

Average 0.083

L. Remarks/Maintenance record: Software 8164.13 upgrade. Annual inspection

08 Dec, 2009  
Date

Deb Kasheer  
Field Inspector's Signature

8 Dec 2009  
Date

Mary Pearson  
Reviewed by



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = TOXL      8164.13.00 06/09  
12/08/2009      08:53

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:53
02 Std. Sol.	0.021	08:54
03 Room Air	0.000	08:55
04 Std. Sol.	0.020	08:55
05 Room Air	0.000	08:56
06 Std. Sol.	0.021	08:57
07 Room Air	0.000	08:57

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

~~Low AC~~ <sup>AC</sup> ~~AC~~ <sup>DK</sup> 0.019 AC

Form 106-I8000

80-003065  
P. 3 of 9 An

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = TOXL      8164.13.00 06/09  
12/08/2009      09:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:09
02 Std. Sol.	0.284	09:10
03 Room Air	0.000	09:10
04 Std. Sol.	0.287	09:11
05 Room Air	0.000	09:11
06 Std. Sol.	0.288	09:12
07 Room Air	0.000	09:13

08 Sim Temp = 34.0°C

Simul Ser No = 00000

Std Sol No = 000

County = 08

Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test 0.285 AC*

Form 106-18000

80-003065  
A. 4-9 FR

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-003065  
Location = TOXL                    8164.13.00 06/09  
12/08/2009                            09:30

Test	AC	Time
01 Room Air	0.000	09:30
02 *Subject Test	INT*	09:30
03 Room Air	0.000	09:31

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 35  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Interferent test*

Form 106-I8000

80-003065  
P. 5 of 9 JAA

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-003065  
Location = TOXL                    8164.13.00 06/09  
12/08/2009                            09:40

Test	AC	Time
01 Diagnostic	OK	09:41
02 Room Air	RFI*	09:41
03 Room Air	0.000	09:42

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 35  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:  
*RFI test*

Form 106-I8000

*80-003065  
P. 6 of 9 1/2*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = TOXL      8164.13.00 06/09  
12/08/2009      09:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:56
02 Std. Gas	0.083	09:56
03 Room Air	0.000	09:57
04 Std. Gas	0.082	09:57
05 Room Air	0.000	09:58
06 Std. Gas	0.082	09:58
07 Room Air	0.000	09:58

Lot No = 659358  
Cyl No = 35  
Exp Date = 09/03/2011  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check*

Form 106-I8000

80-003065  
A 7/9 FM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = TOXL      8164.13.00 06/09  
12/08/2009      09:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:59
02 Std. Gas	0.083	10:00
03 Room Air	0.000	10:00
04 Std. Gas	0.083	10:01
05 Room Air	0.000	10:01
06 Std. Gas	0.083	10:01
07 Room Air	0.000	10:02

Lot No = 659358  
Cyl No = 35  
Exp Date = 09/03/2011  
County = 08      Oper No. = 777777

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check*

Form 106-I8000

80-003065  
P. Pef 9 pm



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003065  
Location = TOXL      8164.13.00 06/09  
12/08/2009      10:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:03
02 Std. Gas	0.083	10:03
03 Room Air	0.000	10:04
04 Std. Gas	0.082	10:04
05 Room Air	0.000	10:05
06 Std. Gas	0.083	10:05
07 Room Air	0.000	10:05

Lot No = 659358  
Cyl No = 35  
Exp Date = 09/03/2011  
County = 08      Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

*Calibration check*

Form 106-I8000

80-003065  
P. 9/29/11