

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
12/09/2009 11:08

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:08
02 Std. Gas	0.083	11:09
03 Room Air	0.000	11:09
04 Std. Gas	0.083	11:10
05 Room Air	0.000	11:10
06 Std. Gas	0.083	11:10
07 Room Air	0.000	11:11

Lot No = 659358
Cyl No = 43
Exp Date = 09/03/2011
County = 22 Oper No. = 119404



Operator Signature
DAVID R. FELAND

Remarks:

Form 106-I8000

80-004186
P. 3 of 4 FL

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
12/09/2009 11:32

Test	AC	Time
01 Diagnostic	OK	11:35
02 Room Air	0.000	11:35
03 *Subject Test	RFI*	11:35
04 Room Air	0.000	11:36

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, RFI TEST RFI
Sub DOB = 01/01/1967
Sub Sex = Male Weight = 200
Test = DUI Cit = 00990
Dr. Lic. = ND/000891
Lot No = 659358
Cyl No = 43
Expiration Date = 09/03/2011
County = 22 Oper No. = 119404

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
DAVID R. FELAND

Remarks:

Form 106-I8000

80-004186
p. 4 of 4