

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David Feland

Location: Kidder SO.

Intoxilyzer Serial No.: 80-004186

Lot Number: 659357

Cylinder Number: 37

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.082 %

Test 2: 0.082 %

Test 3: 0.082 %

(ATTACH TEST RECORD)

David Feland
Chemical Test Operator Signature

09 Dec 09
Date

Deb Keshner
Reviewed By

10 Dec 09
Date

80-004186
P. 1 of 2


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004186
Location = STEE 8164.13.00 06/09
12/09/2009 11:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:01
02 Std. Gas	0.082	11:01
03 Room Air	0.000	11:02
04 Std. Gas	0.082	11:02
05 Room Air	0.000	11:03
06 Std. Gas	0.082	11:03
07 Room Air	0.000	11:04

Lot No = 659358
Cyl No = 43
Exp Date = 09/03/2011
County = 22 Oper No. = 119404


Operator Signature
DAVID R. FELAND

Remarks:

Form 106-I8000

80-004186
P. 2 of 2 FM