

INTOXILYZER® 5000 G AND 8000
INSTALLATION AND REPAIR CHECKOUT

Serial No.: 80-003058 Instrument Location: Ward County Jail

Reason for Install/Repair: Install After Lab Repair/Inspection
 Other (specify) _____

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer® KB.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Ethanol Breath Standard: Connected Installed.
(Include Form 105-G if New Cylinder Installed.)
- F. Review/Enter Gas Setup (Level 1, Function S). (Intoxilyzer® 8000 Only.)
- G. Print Test (Level 1, Function P). (Attach Test Record.)
- H. ACA Test (Level 1, Function C). (Attach Test Record.)
- I. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- J. Repair and/or Maintenance Performed: _____

12/10/2009
Date
24 Dec 2009
Date

Patrick C. Hudson
Field Inspector's Signature

Reviewed By

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = WARD 8164.13.00 06/09
12/10/2009 14:34

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659358
Standard Cyl #? 32
Standard Expiration? 09/03/2011
Oper No? 049509

Flow Cal. Date: 04/07/2008
Slope 698
Intercept -678970

IR Calibration Date: 06/05/2008
 3um 9um

0th Coef(*100): -18867 -13178
1st Coef(*100): 274682 135660
2nd Coef(*100): 1139 678
H2O adj(mg/l*10k): 434 216

***** Printer Test End *****



Operator Signature
PATRICK C. HUDSON

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = WARD 8164.13.00 06/09
12/10/2009 14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:36
02 Std. Gas	0.082	14:37
03 Room Air	0.000	14:37
04 Std. Gas	0.081	14:37
05 Room Air	0.000	14:38
06 Std. Gas	0.082	14:38
07 Room Air	0.000	14:39

Lot No = 659358
Cyl No = 32
Exp Date = 09/03/2011
County = 51 Oper No. = 049509



Operator Signature
PATRICK C. HUDSON

Remarks:

Form 106-I8000

80-003058
R 3 of 4 PA

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = WARD 8164.13.00 06/09
12/10/2009 14:42

Test	AC	Time
01 Diagnostic	OK	14:44
02 Room Air	RFI*	14:44
03 Room Air	0.000	14:45

*Invalid Test
Inhibited - RFI

Sub Name = RFI, TEST NA
Sub DOB = 01/01/1978
Sub Sex = Male Weight = NA
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 659358
Cyl No = 32
Expiration Date = 09/03/2011
County = 51 Oper No. = 049509

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
PATRICK C. HUDSON

Remarks:

Form 106-I8000

80-003058
P. Hudson