



OFFICE OF ATTORNEY GENERAL
 Crime Laboratory Division
 2641 East Main Avenue
 P.O. Box 937
 Bismarck, North Dakota 58502-0937

DEC 14 REC'D

TEL 701-328-6159

ETHANOL BREATH STANDARD CYLINDER INSTALLATION FORM

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Patrick C Hudson

Location: Ward County Jail

Intoxilyzer Serial No.: 80-003058

Lot Number: 659358

Cylinder Number: 32

Expiration Date: 09/03/2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.083 %

Test 2: 0.083 %

Test 3: 0.082 %

(ATTACH TEST RECORD)

Patrick C Hudson
 Chemical Test Operator Signature

12/10/2009
 Date Cylinder Changed

Deb Keshner

24 Dec 2009

Form 105-G
 Toxicology Laboratory
 Crime Laboratory Division

80-003058
 P.1 of 2
 JFH

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003058
Location = TOXL 8164.13.00 06/09
12/10/2009 14:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:27
02 Std. Gas	0.083	14:27
03 Room Air	0.000	14:28
04 Std. Gas	0.083	14:28
05 Room Air	0.000	14:29
06 Std. Gas	0.082	14:29
07 Room Air	0.000	14:29

Lot No = 659358

Cyl No = 32

Exp Date = 09/03/2011

County = 51

Oper No. = 049509



Operator Signature
PATRICK C. HUDSON

Remarks:

Form 106-I8000

80-003058
P. J. J. Jan