

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Robert M. Melby

Location: Divide County SO (CRDS)

Intoxilyzer Serial No.: 80-003065

Lot Number: 659358

Cylinder Number: 6

Expiration Date: 09/03/11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %
RM

Test 2: 0. 082 %

Test 3: 0. 081 %
RM

(ATTACH TEST RECORD)

Robert M. Melby
Chemical Test Operator Signature

12/10/09
Date

Deb Kashner
Reviewed By

15 Dec 09
Date

80-003065
P. 1 of 2 RM


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003065
Location = CROS 8164.13.00 06/09
12/10/2009 14:36

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:37
02 Std. Gas	0.082	14:37
03 Room Air	0.000	14:38
04 Std. Gas	0.082	14:38
05 Room Air	0.000	14:38
06 Std. Gas	0.081	14:39
07 Room Air	0.000	14:39

Lot No = 659358
Cyl No = 6
Exp Date = 09/03/2011
County = 12 Oper No. = 100011




Operator Signature
ROBERT M. MELBY

Remarks:

Test OK

Form 106-18000

The undersigned, having custody of the original record, certifies that the information hereon is a true and correct copy of the original document maintained as part of the activities of this office.

 12/11/09
Signature Date

80-003065
P. 2 of 2