

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Sgt. Joe Knowski

Location: Devils Lake LEC.

Intoxilyzer Serial No.: 80-004193

Lot Number: 059358

Cylinder Number: 20

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %

Test 2: 0. 081 %

Test 3: 0. 081 %

(ATTACH TEST RECORD)

Joe Knowski
Chemical Test Operator Signature

12-10-2009
Date

Deb Keshur
Reviewed By

30 Dec 2009
Date

80-004193
P. 1 of 2
JAN


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004193
Location = DEVL 8164.13.00 06/09
12/10/2009 20:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:08
02 Std. Gas	0.081	20:08
03 Room Air	0.000	20:08
04 Std. Gas	0.081	20:09
05 Room Air	0.000	20:09
06 Std. Gas	0.081	20:10
07 Room Air	0.000	20:10

Lot No = 659358
Cyl No = 20
Exp Date = 09/03/2011
County = 36 Oper No. = 109208



Operator Signature
JOSEPH J. KNOWSKI

Remarks:

Form 106-I8000

80-004193
P. 2 of 2 JKM