

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = WFPD 8164.13.00 06/09
12/11/2009 15:02

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-|=|
ABCDEFGHIJKLMNopqrstuvwxyz!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659357
Standard Cyl #? 10
Standard Expiration? 09/03/2011
Oper No? 049524

Flow Cal. Date: 04/07/2008
Slope 713
Intercept -692349

IR Calibration Date: 06/04/2008
 3um 9um

0th Coef(*100): -15808 -12583
1st Coef(*100): 279401 139580
2nd Coef(*100): 1046 760
H2O adj(mg/l*10k): 478 208

***** Printer Test End *****



Operator Signature
TROY A. HISCHER

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = WFPD 8164.13.00 06/09
12/11/2009 15:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:08
02 Std. Gas	0.081	15:08
03 Room Air	0.000	15:09
04 Std. Gas	0.081	15:09
05 Room Air	0.000	15:10
06 Std. Gas	0.081	15:10
07 Room Air	0.000	15:11

Lot No = 659357
Cyl No = 10
Exp Date = 09/03/2011
County = 09 Oper No. = 049524


Operator Signature
TROY A. HISCHER

Remarks:

Form 106-I8000

80-003059
P. 3 of 4 JEM

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003059
Location = WFPD 8164.13.00 06/09
12/11/2009 15:06

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	15:07
02 Room Air	0.000	15:07

*Invalid Test
Inhibited - RFI

Lot No = 659357
Cyl No = 10
Exp Date = 09/03/2011
County = 09 Oper No. = 049524

 RFI

Operator Signature
TROY A. HISCHER

Remarks:

Form 106-18000

80-003059
P. 4 of 4 JAM