

### ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Troy A. Hischer

Location: West Fargo PD

Intoxilyzer Serial No.: 80-003059

Lot Number: 659357

Cylinder Number: 10

Expiration Date: 9-03-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 081 %

Test 2: 0. 082 %

Test 3: 0. 081 %

(ATTACH TEST RECORD)

Troy A. Hischer  
Chemical Test Operator Signature

12-11-09  
Date

Deb Kashner  
Reviewed By

24 Dec 2009  
Date

80-003059  
P. 1 of 2  
FH

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-003059  
Location = WFPD      8164.13.00 06/09  
12/11/2009      15:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:03
02 Std. Gas	0.081	15:03
03 Room Air	0.000	15:04
04 Std. Gas	0.082	15:04
05 Room Air	0.000	15:05
06 Std. Gas	0.081	15:05
07 Room Air	0.000	15:06

Lot No = 659357  
Cyl No = 10  
Exp Date = 09/03/2011  
County = 09      Oper No. = 049524

Troy A. Hischer ISF ACA  
Operator Signature  
TROY A. HISCHER

Remarks:

Form 106-I8000

80-003059  
P. 2 of 2  
1/11/11