

OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION
Toxicology Section/Breath Alcohol Program

Form: BrF-105-G

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: David M Wallace

Location: Dickinson IEC

Intoxilyzer Serial No.: 80-003066

Lot Number: 459354

Cylinder Number: 23

Expiration Date: 9-3-2011

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 082 %

Test 2: 0. 082 %

Test 3: 0. 084 %

(ATTACH TEST RECORD)

David M Wallace
Chemical Test Operator Signature

1-15-10 12/11/2009 DK
Date

Deb Kashur
Reviewed By

22 Jan 2010
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-003066
Location = DKPD 8164.13.00 06/09
12/11/2009 15:00

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:01
02 Std. Gas	0.082	15:01
03 Room Air	0.000	15:02
04 Std. Gas	0.082	15:02
05 Room Air	0.000	15:03
06 Std. Gas	0.084	15:03
07 Room Air	0.000	15:04

Lot No = 659358
Cyl No = 23
Exp Date = 09/03/2011
County = 45 Oper No. = 068332


Operator Signature
DAVID WALLACE

Remarks:

Form 106-I8000

RCH

80-003066
P. 242 AM