



## Intoxilyzer 8000 Initial Inspection

<b>Serial Number</b>	<b>80-004194</b>
----------------------	------------------

<input checked="" type="checkbox"/> A. Power On	<input checked="" type="checkbox"/> 5 min. countdown	<input checked="" type="checkbox"/> 1 min.												
<input checked="" type="checkbox"/> B. Security Checks.	<input checked="" type="checkbox"/> Level 1	<input checked="" type="checkbox"/> Level 2 <input checked="" type="checkbox"/> Level 3												
<input checked="" type="checkbox"/> C. Gas Cylinder Setup														
<input checked="" type="checkbox"/> D. Printer. Level 1 Func P	<input checked="" type="checkbox"/> Paper Feed	<input checked="" type="checkbox"/> Internal <input checked="" type="checkbox"/> External												
<input checked="" type="checkbox"/> E. Display is working.														
<input checked="" type="checkbox"/> F. Configure Start Test Switch. Level 3 Func S Sub T														
<input checked="" type="checkbox"/> G. Setup Preliminary Data. Level 2 Func E														
<input checked="" type="checkbox"/> H. Configure General Setup. Level 3 Func S Sub G.														
<input checked="" type="checkbox"/> I. Configure Port. Level 3 Func S Sub N														
<input checked="" type="checkbox"/> J. Check Camera. ABA Mode	<input checked="" type="checkbox"/> Oper ID	<input checked="" type="checkbox"/> ND Lic. <input checked="" type="checkbox"/> Hot Key "R"												
<input checked="" type="checkbox"/> K. Diagnostic Test. Level 2 Func D														
<input checked="" type="checkbox"/> L. Check Ser Num / Version Level 2 Func V	80-004194 Serial No. 8164.13 Version 1362.14 Display 7 CPLD													
<input checked="" type="checkbox"/> M. Detector Voltages. Level 3 Func D Sub D	<table style="margin-left: 20px;"> <tr><td>3</td><td><u>13109</u></td><td>s</td><td><u>0</u></td><td>C</td><td><u>47.00</u></td></tr> <tr><td>9</td><td><u>13020</u></td><td>s</td><td><u>4</u></td><td>B</td><td><u>44</u> <u>0</u> RFI Ind.</td></tr> </table>	3	<u>13109</u>	s	<u>0</u>	C	<u>47.00</u>	9	<u>13020</u>	s	<u>4</u>	B	<u>44</u> <u>0</u> RFI Ind.	
3	<u>13109</u>	s	<u>0</u>	C	<u>47.00</u>									
9	<u>13020</u>	s	<u>4</u>	B	<u>44</u> <u>0</u> RFI Ind.									
<input checked="" type="checkbox"/> N. Hot Keys	<table style="margin-left: 20px;"> <tr><td><input checked="" type="checkbox"/></td><td>F5 Tone</td></tr> <tr><td><input checked="" type="checkbox"/></td><td>F6 Green LED</td></tr> <tr><td><input checked="" type="checkbox"/></td><td>F7 Red LED</td></tr> <tr><td><input checked="" type="checkbox"/></td><td>F9 Fan</td></tr> <tr><td><input checked="" type="checkbox"/></td><td>F10 Pump</td></tr> <tr><td><input checked="" type="checkbox"/></td><td>F11 Sol Output</td></tr> </table>	<input checked="" type="checkbox"/>	F5 Tone	<input checked="" type="checkbox"/>	F6 Green LED	<input checked="" type="checkbox"/>	F7 Red LED	<input checked="" type="checkbox"/>	F9 Fan	<input checked="" type="checkbox"/>	F10 Pump	<input checked="" type="checkbox"/>	F11 Sol Output	
<input checked="" type="checkbox"/>	F5 Tone													
<input checked="" type="checkbox"/>	F6 Green LED													
<input checked="" type="checkbox"/>	F7 Red LED													
<input checked="" type="checkbox"/>	F9 Fan													
<input checked="" type="checkbox"/>	F10 Pump													
<input checked="" type="checkbox"/>	F11 Sol Output													
<input checked="" type="checkbox"/> O. RFI Dectector														
<input checked="" type="checkbox"/> P. Atmospheric Monitor. Level 3 Func D Sub A.	P <u>1009</u>	CF <u>1.00</u>												

08/09

80-004194  
P. 14/21  
sm

OFFICE OF ATTORNEY GENERAL  
 CRIME LABORATORY DIVISION  
 Toxicology Section/Breath Alcohol Program  
 Instrument Certification and Maintenance Unit

BrF-005-8

- ✓ Q. Flow Monitor. Level 3 Func D Sub F. R 175 F 0.000 L/s
- ✓ R. Temperature Monitor. Level 3 Func D Sub T  
 C 47.0 A 48.0 I 0.0  
 B 44.9 V 0.0
- ✓ S. Gas Tank Monitor. Level 3 Func D Sub G 462 psi Intox 460 psi Reg
- ✓ T. Configure Wet Bath Sim. Level 1 Func S
- ✓ U. Zero Alcohol Test Level 1 Func C  
 DR 3409 Sim SN — Lot — 0.000 Rep Value  
0.500 AC 0.000 AC 0.000 AC
- ✓ V. Low AC Test. Level 1 Func C  
 DR 5144 Sim SN — Lot — 0.019 Rep Value  
0.019 AC 0.019 AC 0.019 AC
- ✓ W. Linearity Test. Level 1 Func C  
 DR 5187 Sim SN — Lot — 0.285 Rep Value  
~~DR 5187~~ 0.281 AC 0.282 AC 0.282 AC
- ✓ X. Interferent Test. Level 1 Func B  
 Acetone DR 3845 Sim SN — Lot — 0.05% Conc  
 ✓ Display ✓ Printout  
 Methanol 310573 Sim SN — Lot — 0.10% Conc  
 ✓ Display ✓ Printout  
 IPA DR 5191 Sim SN — Lot — 0.10% Conc  
 ✓ Display ✓ Printout
- ✓ Y. Ambient Test  
 ✓ Display ✓ Printout
- ✓ Z. Invalid Sample Test  
 ✓ Display ✓ Printout
- ✓ AA. RFI Detector Test. CMS Mode  
 ✓ Display ✓ Printout
- ✓ BB. Configure Dry Gas Standard. Level 1 Func S
- ✓ CC. Stability Test at 0.080AC. Level 3 Func T Sub S  
0.0792 Ave. 0.0009 Std Dev 1.13% Rel Std Dev
- ✓ DD. Maintenance/Repairs Calibration of optical bench and ITP. Software  
upgrade 8/6/09.

11 Dec 2009  
 Date  
15 Dec 2009  
 Date

Deb Kashur  
 Signature  
Mary Pearson  
 Reviewed by

08/09

80-004194  
 P. 2 of 21  
 FM

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004194  
 12/10/2009 12:18:20

Auto Calibration

	<<<<< 3um >>>>>		<<<<< 9um >>>>>	
-----				
Solution = 0.000 g/210L or 0.0000 mg/l, Samples = 4, Discarded = 1				
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.0790	(0.0120)	0.2790	(0.0140)
Sample #2	0.1310	(0.0060)	0.2520	(0.0410)
Sample #3	0.1220	(0.0150)	0.2700	(0.0510)
Sample #4	0.1100	(0.0270)	0.2620	(0.0620)
Avg % Abs	0.1210	(0.0160)	0.2613	(0.0513)
STD DEV	0.0105	(0.0105)	0.0090	(0.0105)
REL STD DEV	8.707	(65.848)	3.451	(20.462)
-----				
Solution = 0.040 g/210L or 0.1905 mg/l, Samples = 4, Discarded = 1				
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	0.8330	(0.0060)	1.6420	(0.0100)
Sample #2	0.8020	(0.0140)	1.6730	(-0.0030)
Sample #3	0.7840	(0.0260)	1.6530	(0.0200)
Sample #4	0.7940	(0.0290)	1.6420	(0.0240)
Avg % Abs	0.7933	(0.0230)	1.6560	(0.0137)
STD DEV	0.0090	(0.0079)	0.0157	(0.0146)
REL STD DEV	1.137	(34.510)	0.949	(106.622)
-----				
Solution = 0.107 g/210L or 0.5095 mg/l, Samples = 4, Discarded = 1				
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	2.0250	(-0.0130)	4.0100	(0.0000)
Sample #2	1.9790	(0.0130)	4.0100	(0.0070)
Sample #3	1.9720	(0.0250)	3.9940	(0.0230)
Sample #4	2.0170	(-0.0020)	4.0220	(0.0110)
Avg % Abs	1.9893	(0.0120)	4.0087	(0.0137)
STD DEV	0.0242	(0.0135)	0.0140	(0.0083)
REL STD DEV	1.217	(112.731)	0.350	(60.927)
-----				
Solution = 0.147 g/210L or 0.7000 mg/l, Samples = 4, Discarded = 1				
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	2.7020	(-0.0230)	5.3920	(-0.0170)
Sample #2	2.7050	(-0.0090)	5.3670	(0.0000)
Sample #3	2.6670	(0.0170)	5.3450	(0.0220)
Sample #4	2.6750	(0.0120)	5.3540	(0.0290)
Avg % Abs	2.6823	(0.0067)	5.3553	(0.0170)
STD DEV	0.0200	(0.0138)	0.0111	(0.0151)
REL STD DEV	0.747	(206.942)	0.207	(89.016)
-----				
Solution = 0.285 g/210L or 1.3571 mg/l, Samples = 4, Discarded = 1				
Sample	% Abs	(% Abs Ref)	% Abs	(% Abs Ref)
Sample #1	5.0640	(0.0180)	9.8800	(0.0110)
Sample #2	5.1080	(0.0300)	9.9400	(0.0280)
Sample #3	5.1590	(0.0150)	9.9950	(0.0160)
Sample #4	5.1180	(0.0410)	9.9960	(0.0300)
Avg % Abs	5.1283	(0.0287)	9.9770	(0.0247)
STD DEV	0.0270	(0.0131)	0.0320	(0.0076)
REL STD DEV	0.527	(45.527)	0.321	(30.697)
-----				

*Calibration  
 Deb Kashner  
 p. 1 of 2*

*80-004194  
 P 3 of 21, Jan*

TOXL  
 Intoxilyzer - Alcohol Analyzer  
 Model 8000 SN 80-004194  
 12/10/2009 12:18:20

Auto Calibration

```

<<<<<      3um      >>>>>
-----
Zero Order Coef   -308.96
First Order Coef  2742.94
Second Order Coef  -7.15
  
```

```

<<<<<      9um      >>>>>
-----
                -344.19
                1339.69
                5.52
  
```

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0005
0.040	0.039	0.0009
0.107	0.108	-0.0005
0.147	0.147	0.0001
0.285	0.285	0.0000

Act (g/210L)	Fit (g/210L)	Residual (g/210L)
0.000	0.000	-0.0001
0.040	0.040	0.0003
0.107	0.107	-0.0004
0.147	0.147	0.0002
0.285	0.285	0.0000

```

<<<<<      3um      >>>>>      <<<<<      9um      >>>>>
-----
Solution = 0.080 g/210L or 0.3810 mg/l, Samples = 4, Discarded = 1
Sample
Sample #1          3308.00          3191.00
Sample #2          3311.00          3151.00
Sample #3          3273.00          3149.00
Sample #4          3372.00          3193.00
Avg                3318.6667          3164.3333
STD DEV            49.9433           24.8462
REL STD DEV        1.505           0.785
H2O adjust (mg/l*10k) 491           645
  
```

Atmospheric Pressure = 1009

\*\*\*\*\*CALIBRATION SUCCESSFUL\*\*\*\*\*

*Calibration  
 Deb Kashner  
 p. 2 of 2*

80-004194  
 P. 4 of 21 JKA

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004194  
Location = TOXL                8164.13.00 06/09  
12/10/2009                        13:47  
\*\*Internal Test Procedure Calibration\*\*  
Target: 080 ReloadOffset:5989

*ITP test (calibration)  
Deb Kashner  
p. 1 of 1*

*80-004194  
P. 5 of 21 1/22*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004194  
Location = TOXL                 8164.13.00 06/09  
12/10/2009                        13:49

ITP CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:49
02 ITP Check	0.080	13:50
03 Room Air	0.000	13:50

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

ITP check

Form 106-I8000

*p. 1061*

*80-004194  
P. 6421 FM*

Test \*\*\*\*\*

12 char/line  
abcdefghijkl  
ABCDEFGHIJKL

16 char per line  
abcdefghijklmnop  
ABCDEFGHIJKLMNPO

21 chars per line  
abcdefghijklmnopqrstu  
ABCDEFGHIJKLMNOPQRSTU

24 characters per line  
abcdefghijklmnopqrstuwx  
ABCDEFGHIJKLMNOPQRSTUWX

32 characters per line  
abcdefghijklmnopqrstuwx123456  
ABCDEFGHIJKLMNOPQRSTUWXYZ123456

42 characters per line  
abcdefghijklmnopqrstuwx1234567890abcdef  
ABCDEFGHIJKLMNOPQRSTUWXYZ1234567890ABCDEF

Current Instrument Setup

Data Entry Mode: Enabled  
Start Test Sequence: DABACABA  
Display Prelim Rslt? Yes  
Display Third Digit? Yes  
Inhib Printer(Y/N)? No  
Display Volume? No  
Disable On Memfull? Yes  
# of Print Copies? 1  
Select Std (D/W/L)? Dry  
Standard Value? 0.060  
Standard Lot #? 659359  
Standard Cyl #? 54  
Standard Expiration? 09/03/2011  
Oper No? 77777

Flow Cal. Date: 06/23/2009  
Slope: 661  
Intercept: -522430

IR Calibration Date: 12/10/2009  
Sun Sun

0th Coef(\*100): -30896 -34419  
1st Coef(\*100): 274234 133965  
2nd Coef(\*100): -715 552  
H2O adj.(mg/l\*10K): 491 645

\*\*\*\*\* Test End \*\*\*\*\*

Deb Kasher  
Operator Signature  
DEB KASHUR

Remarks:  
Print Test

Form 106-18000

86-009194  
P. 7/21/11

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004194  
Location = TOXL                    8164.13.00 06/09  
12/10/2009                            13:54

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer (Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            659358  
Standard Cyl #?            54  
Standard Expiration?        09/03/2011  
Oper No?                    777777

Flow Cal. Date:            06/23/2009  
Slope                        681  
Intercept                   -622430

IR Calibration Date:        12/10/2009  
                                    3um                            9um

-----  
0th Coef(\*100):            -30896                        -34419  
1st Coef(\*100):            274294                        133969  
2nd Coef(\*100):            -715                            552  
H2O adj(mg/l\*10k):        491                            645

\*\*\*\*\* Printer Test End \*\*\*\*\*

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

Print test

Form 106-I8000

80-004194  
P. P. 11/11/12



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004194  
Location = TOXL                    8164.13.00 06/09  
12/10/2009                            14:43

Test	AC	Time
01 Room Air	0.000	14:44
02 *Subject Test	REF*	14:44
03 Room Air	0.000	14:45

\*Subject Test Refused

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 54  
Expiration Date = 09/03/2011  
County = 08                          Oper No. = 777777

Deb Kashur

Operator Signature  
DEB KASHUR

Remarks:

*Camera / Refusal check*

Form 106-I8000

80-004194  
P. 9/21 PM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004194  
Location = TOXL                    8164.13.00 06/09  
12/10/2009                            14:46

DIAGNOSTIC

Voltage/Current Test	Pass
RAM Test	Pass
EEPROM Checksum Test	Pass
Real Time Clock Test	Pass
DSP Test	Pass
Analytical Stability Test	Pass
Modem Test	Pass
Temp Regulation Test	Pass

Oper No. = 77777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Diagnostic test*

Form 106-I8000

80-004194  
P. 10/21/09

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004194  
Location = TOXL                8164.13.00 06/09  
12/10/2009                        14:52

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:52
02 Std. Sol.	0.000	14:53
03 Room Air	0.000	14:54
04 Std. Sol.	0.000	14:54
05 Room Air	0.000	14:55
06 Std. Sol.	0.000	14:55
07 Room Air	0.000	14:56

08 Sim Temp = 34.0°C

Simul Ser No = 000

Std Sol No = 0000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*zero alcohol 0.000 AC*

Form 106-18000

*80-004194  
P. 11 of 21 VSH*

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004194  
Location = TOXL                 8164.13.00 06/09  
12/10/2009                        15:08

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:08
02 Std. Sol.	0.019	15:18
03 Room Air	0.000	15:19
04 Std. Sol.	0.019	15:20
05 Room Air	0.000	15:20
06 Std. Sol.	0.019	15:21
07 Room Air	0.000	15:21

08 Sim Temp = 34.0°C

Simul Ser No = 000  
Std Sol No = 0000  
County = 08

Oper No. = 777777

Deb Kashur  
Operator Signature  
DEB KASHUR

Remarks:

Low AC      0.019 AC

Form 106-18000

80-004194  
P. 12/21/09

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004194  
Location = TOXL      8164.13.00 06/09  
12/10/2009      15:27

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:27
02 Std. Sol.	0.281	15:28
03 Room Air	0.000	15:29
04 Std. Sol.	0.282	15:29
05 Room Air	0.000	15:30
06 Std. Sol.	0.282	15:31
07 Room Air	0.000	15:31

08 Sim Temp = 34.0°C

Simul Ser No = 000

Std Sol No = 0000

County = 08

Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*Linearity test*

Form 106-18000

80-004194  
P. 13 of 21 pm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004194  
Location = TOXL      8164.13.00 06/09  
12/11/2009      07:19

Test	AC	Time
01 Room Air	0.000	07:19
02 *Subject Test	INT*	07:20
03 Room Air	0.000	07:20

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 54  
Expiration Date = 09/03/2011  
County = 08      Oper No. = 777777

*Deb Kasher*

Operator Signature  
DEB KASHUR

Remarks:

*acetone interferent*

Form 106-I8000

80-004194  
P. 14 of 21 pages

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004194  
Location = TOXL                              8164.13.00 06/09  
12/11/2009                                      07:21

Test	AC	Time
01 Room Air	0.000	07:22
02 *Subject Test	INT*	07:22
03 Room Air	0.000	07:23

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                              Weight = 150  
Test = DUI                                      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 54  
Expiration Date = 09/03/2011  
County = 08                                      Oper No. = 777777

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*Methanol interferent*

Form 106-I8000

80-004194  
P. 15 of 21 JEM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004194  
Location = TOXL                    8164.13.00 06/09  
12/11/2009                            07:26

Test	AC	Time
01 Room Air	0.000	07:27
02 *Subject Test	INT*	07:27
03 Room Air	0.000	07:28

\*Invalid Test  
Interferent Detected

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                    Weight = 150  
Test = DUI                            Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 54  
Expiration Date = 09/03/2011  
County = 08                            Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*IPA interferent*

Form 106-I8000

*80-004194  
P.16 of 21 ycm*



Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004194  
Location = TOXL      8164.13.00 06/09  
12/11/2009      07:43

Test	AC	Time
01 Room Air	AMB*	07:43
02 Room Air	0.000	07:44

\*Invalid Test  
Check Ambient Conditions

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female      Weight = 150  
Test = DUI      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 54  
Expiration Date = 09/03/2011  
County = 08      Oper No. = 777777

*Deb Kashur*

Operator Signature  
DEB KASHUR

Remarks:

*ambient test*

Form 106-I8000

80-004194  
P. 12-11-09

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004194  
Location = TOXL                8164.13.00 06/09  
12/11/2009                        07:48

Test	AC	Time
01 Room Air	0.000	07:49
02 Invalid Sample	X.XXX	07:49
03 Room Air	0.000	07:49

\*Invalid Test - Mouth Alcohol

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                Weight = 150  
Test = DUI                        Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 54  
Expiration Date = 09/03/2011  
County = 08                        Oper No. = 777777

Deb Keshur

Operator Signature  
DEB KASHUR

Remarks:

*Invalid sample test*

Form 106-18000

80-004194  
P. 18 of 21

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
North Dakota Model 8000                      SN 80-004194  
Location = TOXL                              8164.13.00 06/09  
12/11/2009                                      07:52

Test	AC	Time
01 Diagnostic	OK	07:52
02 Room Air	RFI*	07:52
03 Room Air	0.000	07:53

\*Invalid Test  
Inhibited - RFI

Sub Name = DISCOVER, THE SPIRIT  
Sub DOB = 02/01/1992  
Sub Sex = Female                              Weight = 150  
Test = DUI                                      Cit = NA  
Dr. Lic. = ND/DIS921456  
Lot No = 659358  
Cyl No = 54  
Expiration Date = 09/03/2011  
County = 08                                      Oper No. = 777777

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.

Deb Kasher

Operator Signature  
DEB KASHUR

Remarks:

*RFI test*

Form 106-I8000

80-004194  
P. 19 of 21 PM

Intoxilyzer Test Record and Checklist  
 NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer                      Alcohol Analyzer  
 North Dakota Model 8000                      SN 80-004194  
 Location = TOXL                              8164.13.00 06/09  
 12/11/2009                                      07:58

Test	AC	Time
01 Room Air	0.000	07:59
02 Std. Gas	0.077	07:59
03 Room Air	0.000	08:00
04 Std. Gas	0.078	08:00
05 Room Air	0.000	08:00
06 Std. Gas	0.078	08:01
07 Room Air	0.000	08:01
08 Std. Gas	0.078	08:02
09 Room Air	0.000	08:02
10 Std. Gas	0.079	08:03
11 Room Air	0.000	08:03
12 Std. Gas	0.080	08:04
13 Room Air	0.000	08:04
14 Std. Gas	0.080	08:05
15 Room Air	0.000	08:05
16 Std. Gas	0.079	08:05
17 Room Air	0.000	08:06
18 Std. Gas	0.079	08:06
19 Room Air	0.000	08:07
20 Std. Gas	0.080	08:07
21 Room Air	0.000	08:08
22 Std. Gas	0.080	08:08
23 Room Air	0.000	08:09
24 Std. Gas	0.080	08:09
25 Room Air	0.000	08:10
26 Std. Gas	0.079	08:10
27 Room Air	0.000	08:10
28 Std. Gas	0.079	08:11
29 Room Air	0.000	08:11
30 Std. Gas	0.080	08:12
31 Room Air	0.000	08:12
32 Std. Gas	0.080	08:13
33 Room Air	0.000	08:13
34 Std. Gas	0.079	08:13
35 Room Air	0.000	08:14
36 Std. Gas	0.080	08:14
37 Room Air	0.000	08:15
38 Std. Gas	0.079	08:15
39 Room Air	0.000	08:16
40 Std. Gas	0.080	08:16
41 Room Air	0.000	08:17
42 Cal Check Stats:		
43 Average	0.0792	
44 Std Dev	0.0009	
45 Rel Std Dev	1.13%	

*Stability test  
 Deb Keshur  
 p. 1 of 2*

Operator Signature  
DEB KASHUR

Remarks:

Lot #: 659358 Cyl #: 54 Exp date: 03 Sept 09

Form 106-I8000

Stability test  
Deb Kashur  
p. 2 of 2

80-004194  
P. 21 of 21 KH