

**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**

Serial No.: 80-004205 Instrument Location: Cass Co. Jail

Reason for Install/Repair:  Install After Lab Repair/Inspection  
 Other (Specify) \_\_\_\_\_

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer®.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: \_\_\_\_\_

12-11-09  
Date

[Signature]  
Field Inspector's Signature

24 Dec 2009  
Date

[Signature]  
Reviewed By

80-004205  
P. 1 of 4 gm

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004205  
Location = CASS      8164.13.00 06/09  
12/11/2009      13:52

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=  
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^\*()\_+?

Current Instrument Setup


Data Entry Mode:      Enabled  
Start Test Sequence:      DABACABA  
Display Prelim Rslt?      Yes  
Display Third Digit?      Yes  
Inhib Printer(Y/N)?      No  
Display Volume?      No  
Disable On Memfull?      Yes  
# of Print Copies?      1  
Select Std (D/W/I)?      Dry  
Standard Value?      0.080  
Standard Lot #?      659357  
Standard Cyl #?      40  
Standard Expiration?      09/03/2011  
Oper No?      049524

Flow Cal. Date:      06/04/2009  
Slope      704  
Intercept      -610062

IR Calibration Date:      06/03/2009

	3um	9um
0th Coef(*100):	-16944	-29673
1st Coef(*100):	265939	132317
2nd Coef(*100):	1696	900
H2O adj(mg/l*10k):	490	583

\*\*\*\*\* Printer Test End \*\*\*\*\*

  
Operator Signature  
TROY A. HISCHER

Remarks:

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004205  
Location = CASS      8164.13.00 06/09  
12/11/2009      14:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:04
02 Std. Gas	0.083	14:04
03 Room Air	0.000	14:04
04 Std. Gas	0.083	14:05
05 Room Air	0.000	14:05
06 Std. Gas	0.083	14:06
07 Room Air	0.000	14:06

Lot No = 659357  
Cyl No = 40  
Exp Date = 09/03/2011  
County = 09      Oper No. = 049524

Troy A. Hischer 2<sup>nd</sup> aca.  
Operator Signature  
TROY A. HISCHER

Remarks:

Form 106-I8000

80-004205  
p. 3 of 4 AM

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004205  
Location = CASS      8164.13.00 06/09  
12/11/2009      14:01

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	14:01
02 Room Air	0.000	14:02

\*Invalid Test  
Inhibited - RFI

Lot No = 659357  
Cyl No = 40  
Exp Date = 09/03/2011  
County = 09      Oper No. = 049524

 RFI

Operator Signature  
TROY A. HISCHER

Remarks:

Form 106-I8000

80-004205  
P-4 of 4 RM