

### ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Troy A. Hirschen

Location: Cass Co Jail

Intoxilyzer Serial No.: 80-004205

Lot Number: 659357

Cylinder Number: 40

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0.083 %

Test 2: 0.083 %

Test 3: 0.083 %

(ATTACH TEST RECORD)

Troy Hirschen  
Chemical Test Operator Signature

12-11-09  
Date

Deb Kashner  
Reviewed By

24 Dec 2009  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004205  
Location = CASS      8164.13.00 06/09  
12/11/2009      13:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:57
02 Std. Gas	0.083	13:58
03 Room Air	0.000	13:58
04 Std. Gas	0.083	13:59
05 Room Air	0.000	13:59
06 Std. Gas	0.083	13:59
07 Room Air	0.000	14:00

Lot No = 659357

Cyl No = 40

Exp Date = 09/03/2011

County = 09

Oper No. = 049524

 *Troy Hischer* P. 2 of 2

Operator Signature  
TROY A. HISCHER

Remarks:

Form 106-18000

80-004205  
P. 2 of 2