

BREATH ALCOHOL INSTRUMENT CHAIN OF CUSTODY WORKSHEET

Instrument: ~~80-003061~~ ^{PK} 8000 Serial Number: 80-003061

Transported Out of the Crime Laboratory:

Agency Being Assigned Instrument: _____

Name/Agency of Person Transporting Instrument:

(Print Name) _____ (Print Agency)

Signature of Person Receiving Instrument: _____

Date Receiving Instrument: ____/____/____ From: _____

Or Shipped Via:
USPS _____ UPS _____ FedEx _____ Other _____

Date Sending Instrument: ____/____/____ Air Bill No. _____

Complete Air Bill (Identify Instrument/Serial Number) and attach a copy to this worksheet.

Transported to the Crime Laboratory:

Agency Returning Instrument: CM1

Name/Agency of Person Returning Instrument:

(Print Name) _____ (Print Agency)

Signature of Person Returning Instrument: _____

Date Returning Instrument: ____/____/____ Received By: _____

Or Received Via:
USPS _____ UPS _____ FedEx Other _____

Date Receiving Instrument: 13 / Dec / 2010 Air Bill No. _____

Attach shipping paperwork (if any) to this worksheet.

Instrument Transfer Information Entered Into the Equipment Database:

Date of Action 13 / Dec / 2010 By: Deb Kashner

80-003061
P. 110-112
PK

INTOX
PBT
TOXITEST

DATE:	10/29/10
RECEIVED VIA:	FEDX

CMI INC.
316 E. 9th St., Owensboro, KY 42303
270-685-6200 — Fax 270-685-6288
1-800-835-0690

SERVICE WORK ORDER

325352

B
I
L
L
T
O

ATTN:		
CITY	STATE	ZIP
CUSTOMER NO.	SRS 013	
PHONE		

S
H
I
P
T
O

Office of Attorney General		
ATTN: Deb Kachur		
2641 Main Avenue		
Crime Lab Div.		
CITY	STATE	ZIP
Rismark	ND	58501
PHONE	701-328-6159	

MODEL #	SN	DATE RCVD.	P.O. #	REQ.	BILL CODE	SERVICE LOCATION CODE
8000	80-003061				C	
PROD. CODE	EXTRA PARTS RCVD. PC, BT, SA, box, battery					

DESCRIPTION OF PROBLEM: Voltage errors during ~~initial~~ diagnostics

QUOTE WARRANTY SERVICE DEPT. NEW

- PRELIMINARY TEST DISMANTLE, INSPECT, DIAGNOSTIC TEST
- POWER SUPPLY TEST, INSPECT REPLACE
- MAIN PWB's - REMOVE & CLEAN, REINSTALL, TEST
- PWB's DIAGNOSE, REPAIR REPLACE
- PWB's - SURFACE CLEAN IN PLACE
- FILTER WHEEL - REMOVE, CLEAN, REINSTALL, CHECK CELL INTERIOR
- CHOPPER WHEEL MOTOR TEST & INSPECT REPLACE
- CELL - REMOVE, CLEAN, REINSTALL, REPLACE LENSES/GASKETS REQ'D
- IR LIGHT SOURCE REMOVE, REALIGN REPLACE
- PRINTER ASSY. DISASSEMBLE CLEAN & LUBRICATE
- REMOVE & REPLACE BELT REALIGNMENT
- REPLACE TEST PRINTER FUNCTIONS
- BATTERIES - REPLACE CPU RAM
- REPLACE PARTS REASSEMBLE CALIBRATION FINAL TEST
- MISC.
- PBT PRELIMINARY TESTING - PRESSURE, BATTERY, ZERO SET
- DISASSEMBLE REPLACE PARTS REASSEMBLE
- TEST CALIBRATE FINAL TEST
- TOXITEST INCOMING LEAK TEST VISUAL/MECH. INSPECTION
- TEST THERMOMETER & THERMOSTAT DISASSEMBLE
- REPLACE PARTS REASSEMBLE OUTGOING & LEAK TEST
- FINAL TEST
- MISC.

PARTS USED			
QTY	ID NO.	DESCRIPTION	CODE
1	013010	Battery (renew)	

COMMENTS: Replaced parts listed

Seal # 03810

HOURS:	2.5 HR
TECH.	K. A. [initials]
DATE	11/17/10
CAL. TECH.	JK
DATE	11/23/10

SHIPPING DATE/CHARGES

80-003061
P-111 of 112
JM

Certificate of Calibration

Intoxilyzer® 8000

*This is to certify that the calibration of **Intoxilyzer**® serial number 80-003061, manufactured by CMI Inc., a subsidiary of MPD, Inc. of Owensboro, Kentucky, was tested and found to be in compliance with the National Highway Traffic Safety Administration Standard for Devices to Measure Breath Alcohol (F.R., No. 179 48705-48710 Sept. 17, 1993) and with International Recommendation OIML R126 Edition 1998 (E). Calibration solutions are traceable to NIST standard material 1828.*

Date 11/23/10

Signed

Kristi Rolando

Technician

EMI INC.

80-003061
p. 112 of 112
PR

316 East 9th Street
Owensboro, KY 42303
USA

Part No. 650519 CMI 7/9/09