

ETHANOL BREATH STANDARD CYLINDER REPORT

(Only Chemical Test Operators may change the Ethanol Breath Standard and complete this form.)

Chemical Test Operator Name: Barry Vennette

Location: MOBL

Intoxilyzer Serial No.: 86-004198

Lot Number: 659358

Cylinder Number: 27

Expiration Date: 9-3-11

Instrument Test Results (Report to 3 Digits; Example, 0.081%)

Test 1: 0. 080 %

Test 2: 0. 080 %

Test 3: 0. 079 %

(ATTACH TEST RECORD)

Barry Vennette
Chemical Test Operator Signature

12-14-09
Date

Deb Kashner
Reviewed By

16-Dec 09
Date

80-004198
P. 1/2 Jm


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = MOHL 8164.13.00 06/09
12/14/2009 20:35

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:36
02 Std. Gas	0.080	20:36
03 Room Air	0.000	20:37
04 Std. Gas	0.080	20:37
05 Room Air	0.000	20:38
06 Std. Gas	0.079	20:38
07 Room Air	0.000	20:39

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 039909



Operator Signature
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004198
P. 2 of 2 FM