

INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

Serial No.: 80-004198 Instrument Location: Mohl.

Reason for Install/Repair: _____ Install After Lab Repair/Inspection
X Other (Specify) New Machine

Check When Done:

- A. Surge Protector Installed/Properly Grounded.
- B. Telephone Line Connected to Intoxilyzer®.
- C. Breath Tube Heated.
- D. Review/Enter Preliminary Data Entry (Level 2, Function E).
- E. Review/Enter Gas Setup (Level 1, Function S).
- F. Print Test (Level 1, Function P). (Attach Test Record.)
- G. ACA Test (Level 1, Function C). (Attach Test Record.)
- H. RFI Test (Use CMS Mode-First Room Air). (Attach Test Record.)
- I. Repair and/or Maintenance Performed: _____

12-14-09
Date

[Signature]
Field Inspector's Signature

16 Dec 2009
Date

Deb Kashner
Reviewed By

80-004198
P-1-4-11

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = MOHL 8164.13.00 06/09
12/14/2009 20:25

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 659358
Standard Cyl #? 27
Standard Expiration? 09/03/2011
Oper No? 039909

Flow Cal. Date: 06/04/2009
Slope 671
Intercept -536290

IR Calibration Date: 12/09/2009
 3um 9um

	3um	9um
0th Coef(*100):	-31070	-34849
1st Coef(*100):	268664	135239
2nd Coef(*100):	-643	440
H2O adj(mg/l*10k):	570	615

***** Printer Test End *****



Operator Signature
BARRY A. VANNATTA

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = MOHL 8164.13.00 06/09
12/14/2009 20:26

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	20:27
02 Std. Gas	0.079	20:27
03 Room Air	0.000	20:28
04 Std. Gas	0.079	20:28
05 Room Air	0.000	20:29
06 Std. Gas	0.079	20:29
07 Room Air	0.000	20:30

Lot No = 659358
Cyl No = 27
Exp Date = 09/03/2011
County = 38 Oper No. = 039909



Operator Signature
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004198
P. 3 of 4

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004198
Location = MOHL 8164.13.00 06/09
12/14/2009 20:31

Test	AC	Time
01 Diagnostic	OK	20:33
02 Room Air	RFI*	20:33
03 Room Air	0.000	20:34

*Invalid Test
Inhibited - RFI

Sub Name = TEST, SUBJECT NA
Sub DOB = 01/17/1966
Sub Sex = Male Weight = 100
Test = OTH Cit = NA
Dr. Lic. = ND/NA
Lot No = 659358
Cyl No = 27
Expiration Date = 09/03/2011
County = 38 Oper No. = 039909

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
BARRY A. VANNATTA

Remarks:

Form 106-I8000

80-004198
P. 4 of 4