



ETHANOL GAS STANDARD CYLINDER REPORT
 OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION
 Toxicology Section/Breath Alcohol Program
 SFN 59282 (10-2013)

Chemical Test Operator Name (Print) <u>Kyle J. Johnson #664</u>		
Location <u>800 4th Ave. E. West Fargo, ND</u>	Intoxilyzer® Serial Number <u>80-004944</u>	
Gas Lot Number <u>14214080A6</u>	Gas Cylinder Number <u>59</u>	Gas Expiration Date <u>07/05/2016</u>

Check When Done:

- 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 2. Perform an ACA Test (Level 1, Function C).
 Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.082 AC
 - B. 0.082 AC
 - C. 0.082 AC
- 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <u>[Signature] #664</u>	Date <u>12/14/2014</u> ✓
Reviewed By (Crime Laboratory Use Only) <u>Deb Kasher</u>	Date <u>15 Dec. 2014</u>


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = WFPD 8164.13.00 06/09
12/14/2014 18:51

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:52
02 Std. Gas	0.082	18:52
03 Room Air	0.000	18:52
04 Std. Gas	0.082	18:53
05 Room Air	0.000	18:53
06 Std. Gas	0.082	18:54
07 Room Air	0.000	18:54

Lot No = 14214080A6
Cyl No = 59
Exp Date = 07/05/2016
County = 09 Oper No. = 130781


Operator Signature
KYLE JOHNSON

Remarks: